

Data Validation Report

Project:	PPG Site 156 - Metro Towers				
Laboratory:	aboratory: SGS Accutest, Dayton, NJ				
Laboratory Job No.:	JC19972				
Analysis/Method:	Hexavalent Chromiun Metals by ICP-AES/ S	n SW846 3060A/7196 SW-846 6010C			
Validation Level:	Full (Hexavalent Chro Limited (Metals)	omium)			
Site Location/Address:	PPG Site 156 - Metro	Towers, Jersey City, NJ			
AECOM Project No:	60493065.20				
Prepared by: Dawn Brule /AECOM		Completed on: 06/29/2016			
	k/AECOM	File Name: JC19972_2016-06-29_DV Report-F			
1.54.5 d 54.5 c.5					

Introduction

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP and/or Region 2 validation Standard Operating Procedure(s) (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, and USEPA SW-846 Method 7196A:
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods).
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015)

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.
- JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in

AECOM 2

the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.

RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on May 10, 2016 as part of the sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
156-FB20160510 (Equipment Blank)	JC19972-6	Aqueous	Hexavalent Chromium
156-FB20160510 (Equipment Blank)	JC19972-6A	Aqueous	Metals (Sb, Cr, Ni, V, Tl)
156-MW1R-20160510	JC19972-5	Ground Water	Hexavalent Chromium
156-MW1R-20160510	JC19972-5A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW5R-20160510	JC19972-2	Ground Water	Hexavalent Chromium
156-MW5R-20160510	JC19972-2A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW6-20160510	JC19972-3	Ground Water	Hexavalent Chromium
156-MW6-20160510	JC19972-3A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW6D-20160510 (field duplicate of 156-MW6-20160510)	JC19972-4	Ground Water	Hexavalent Chromium
156-MW6D-20160510 (field duplicate of 156-MW6-20160510)	JC19972-4A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW7-20160510	JC19972-1	Ground Water	Hexavalent Chromium
156-MW7-20160510	JC19972-1A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

MS Results

Sample 156-MW7-20160510 was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The MS recovery

AECOM 3

was 93.3%; which met the quality control criteria of 75-125%. No data qualification was required on the basis of spike recovery.

Metals

Equipment Blank

Nickel and vanadium were detected in the equipment blank associated with the samples in this SDG, at a concentration above the MDL, but below the RL.

Nickel was detected at less than three times the EB in samples 156-MW7-20160510, 156-MW5R-20160510, and 156-MW1R-20160510; therefore, the results were negated at the RL.

Vanadium was detected at less than three times the EB in sample 156-MW5R-20160510; therefore, the result was negated at the RL

Nickel and vanadium were also estimated (J) in additional samples since the concentrations were detected at greater than three times, but less than ten times the EB.

Sample Results

Reported results (flagged B by the laboratory) that were less than the RL, but greater than or equal to the MDL, are approximate values and have been qualified as estimated (J).

Data Quality and Usability

In general, these data appear to be valid and may be used for decision-making purposes. No data were rejected. Qualified results, if applicable, are discussed in attachments A and B below.

Sample results considered to be negated due to blank contamination are usable as nondetect results at the RL.

Sample results reported between the MDL and RL are usable as estimated values.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hit List(s)

AECOM Page 1 of 3

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers

Sampling Date May 10, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC19972 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB20160510

Field Sample ID	Lab Sample ID	Analyte	Method Blank (mg/l)	,	Validation Sample Result (mg/l)	RL (mg/l)	Assurance	NJDEP Validation Footnote
156-MW1R-20160510	JC19972-5	CHROMIUM (HEXAVALENT)	U	0.0072B	0.0072	0.010	Qualify	1
156-MW6-20160510	JC19972-3	CHROMIUM (HEXAVALENT)	U	0.0040B	0.0040	0.010	Qualify	1

Note: A "U" under Method Blank column indicates a nondetect result.

A "U" under the Laboratory Sample Result and Validation Sample Result columns indicates a nondetect result at the RL.

NJDEP Laboratory Footnote

1. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

AECOM Page 2 of 3

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 - Metro Towers

May 10, 2016

Sampling Date Lab Name/ID Accutest, Dayton, NJ

SDG No JC19972 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB20160510

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Laboratory Sample Result (ug/l)	Validation Sample Result (ug/l)	RL (ug/l)	Quality Assurance Decision	NJDEP Validation Footnote
156-MW7-20160510	JC19972-1A	CHROMIUM	U	3.1B	3.1	10	Qualify	3
156-MW5R-20160510	JC19972-2A	CHROMIUM	U	1.5B	1.5	10	Qualify	3
156-MW6-20160510	JC19972-3A	CHROMIUM	U	1.7B	1.7	10	Qualify	3
156-MW6D-20160510	JC19972-4A	CHROMIUM	U	1.7B	1.7	10	Qualify	3
156-MW1R-20160510	JC19972-5A	CHROMIUM	U	4.2B	4.2	10	Qualify	3
156-MW7-20160510	JC19972-1A	NICKEL	U	4.1B	U	10	Negate	1
156-MW5R-20160510	JC19972-2A	NICKEL	U	1.3B	U	10	Negate	1
156-MW6-20160510	JC19972-3A	NICKEL	U	18.0	18.0	10		
156-MW6D-20160510	JC19972-4A	NICKEL	U	16.6	16.6	10	Qualify	2
156-MW1R-20160510	JC19972-5A	NICKEL	U	0.90B	U	10	Negate	1,3
156-FB20160510	JC19972-6A	NICKEL	U	1.7B	1.7	10	Qualify	3
156-MW7-20160510	JC19972-1A	VANADIUM	U	7.5B	7.5	50	Qualify	2,3
156-MW5R-20160510	JC19972-2A	VANADIUM	U	1.4B	U	50	Negate	1,3
156-MW6-20160510	JC19972-3A	VANADIUM	U	17.7B	17.7	50	Qualify	3
156-MW6D-20160510	JC19972-4A	VANADIUM	U	16.3B	16.3	50	Qualify	3
156-MW1R-20160510	JC19972-5A	VANADIUM	U	7.8B	7.8	50	Qualify	2,3
156-FB20160510	JC19972-6A	VANADIUM	U	0.80B	0.8	50	Qualify	3

AECOM Page 3 of 3

Note: A "U" under Method Blank column indicates a nondetect result.

A "U" under the Laboratory Sample Result and Validation Sample Result columns indicates a nondetect result at the RL.

NJDEP Laboratory Footnote

1. The value reported is less than or equal to 3x the value in the trip/field blank. It is the policy of NJDEP-DPFSR to negate the reported value as due to probable foreign contamination unrelated to the actual sample. The end-user, however, is alerted that a reportable quantity of the analyte was detected.

- 2. The value reported is greater than 3x but less than ten (10) the value in the trip/field blank and is considered "real". However, the reported value must be quantitatively qualified "J" due to trip/field blank contamination.
- 3. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

Attachment B

Data Validation Report Form

Client Name: PPG Industries	Project Number: 60493065.20		
Site Location: PPG Site 156 - Metro Towers, Jersey City, NJ	Project Manager: Nan Conrey		
Laboratory: SGS Accutest, Dayton, NJ	Type of Validation: Full		
Laboratory Job No: JC19972	Date Checked: 6/27/2016		
Validator: Dawn Brule	Peer: Mary Kozik		

ITEM	YES	NO	N/A	COMMENTS
Sample results included?	Х			
Reporting Limits met project requirements?	Х			
Field I.D. included?	Х			
Laboratory I.D. included?	Х			
Sample matrix included?	Х			
Sample receipt temperature 2-6°C?	Х			2.7°C
Signed COCs included?	Х			
Date of sample collection included?	Х			
Date of sample digestion included?	Х			
Holding time to digestion met criteria? (Soils -30 days from collection to digestion.)	Х			
Date of analysis included?	Х			
Holding time to analysis met criteria? (Soils -168 hours from digestion to analysis; Aqueous - 24 hours from collection to analysis.	х			
Method reference included?	Х			
Laboratory Case Narrative included?	Х			

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	х			
2) Correlation coefficient of =0.995 (7196A) or =0.999 (7199)	х			
3) Calibrate daily or each time instrument is set up.	Х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	х			
1) %R criteria met? (90 - 110%)	Х			
2) Correct frequency of one per every 10 samples	х			
CCS and QCS from independent source and at mid-level of calibration curve	x			
Calibration Blanks	Х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	х			
2) Absolute value should not exceed MDL.	Х			
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	х			
Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	Х			
1) Eh and pH data was included and plotted for all samples?	х			
Soluble Matrix Spike Data Included in Lab Package?	х			
1) Soluble Matrix %R criteria met? (75-125%R).	х			
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?	х			Spiked at 0.15 mg/L
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	х			
Insoluble Matrix Spike Data Included in Lab Package?			х	NA for aqueous samples
1) Insoluble Matrix %R criteria met? (75-125%R).			х	
2) Was the spike concentration around 400 to 800 mg/Kg?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
Post Digestion Spike			х	NA for aqueous samples
1) Post Digestion Spike %R criteria met? (85-115%R).			х	
Was the spike concentration 40 mg/Kg or twice the sample concentration?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	
Sample Duplicate Data Included in Lab Package?	х			
1) RPD criteria met? (RPD ≤ 20%) if both results are =4x RL or absolute difference ≤ RL if either or both results are <4xRL.	х			
Was a sample duplicate analyzed at the frequency of 1 per batch or 20 samples?	х			
Was a Laboratory Control Sample (LCS) Included in Lab Package?	х			
1) %R criteria met? (80-120%R).	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	х			
Were any Field Duplicate samples submitted with this SDG?	х			
Were Field duplicate RPD criteria met? (RPD≤20% for both sample results >4xRL or absolute difference ≤ RL if either or both results are <4xRL.	х			abs diff <rl< td=""></rl<>
Were all sample quantitation and reporting requirements met?	х			
Were all solid samples reported with percent solids > 50%?			х	NA for aqueous samples
2) Were any samples analyzed or reported with dilutions?		х		
Miscellaneous Items				
1) For soils by 7196A, was the pH within a range of 7.0-8.0?			х	
2) For soils by 7199, was the pH within a range of 9.0-9.5?			х	
3) For aqueous by 7196A, was the pH with a range of 1.5-2.5?	х			
4) For soils (3060A), was the digestion temperature 90-95C for at least 60 minutes?			х	
5) For 7199, was each sample injected twice and was the RPD <20?			х	

SDG#: JC19972			٦	
Batch: GN45633	x - concentration	y - response		
Cr+6 ICAL 05/10/2016	0	0.001		
Aqueous	0.01	0.011		
(p. 130 of data pkg)	0.05	0.043		
(p. 130 of data pkg)	0.03	0.098		
	0.3	0.090		
	0.5	0.471		
	0.8	0.749		
	1	0.749		
	'	0.54		(p. 130 of data pkg)
AECOM Calculated Intercept	0.0003	OK	Reported intercept	0.0003
AECOM Slope	0.9383	OK	Reported Slope	0.9383
AECOM Calculated r	0.99997	OK	Reported r	0.99997
			·	
LCS calculation	GN45633-B1 pgs	. 126,130		
Background Absorbance	0			
Total absorbance	0.141			
Total absorbance - background	0.141			
Instrument Concentration	0.150			
Sample volume (mL)	50			
Final Volume (mL)	50			
Dilution Factor	11			
AECOM Calculated LCS Result (mg/L)	0.15	OK	Reported Result (mg/L)	0.15
0/ B	nao 126 120			
%R = Found/True*100	pgs. 126,130			
True Value (mg/L)	0.15	01/	D = = = = = 1.07 D	100.0
AECOM Calculated %R	100.0	OK	Reported %R	100.0
MS calculation	GN45633-S2 (JC ²	19912-1) pgs. 12	8, 130	
Background reading	0.004			
Total absorbance	0.135			
Total absorbance - background	0.131			
Instrument Concentration	0.1393			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1			
AECOM Calculated MS Result (mg/L)	0.14	OK, rounding	Reported Result (mg/L)	0.14
%R = Found/True*100	GN45633-S2 (JC	19912-1) pgs. 12	8, 130	
True Value (mg/L)	0.15	, . U		
Native concentration (mg/L)	0			
AECOM%R	93	OK rounding	Reported %R	93
Reporting Limit	[156-MW7-20160	510] (JC19912-1)	pgs. 9, 130	
Low Standard	0.01			
Initial volume (mL)	50			
Final volume (mL)	50			
Percent solids	1			
Dilution Factor	1			
Reporting Limit (mg/L)	0.010	OK	Reported RL (mg/L)=	0.010
-	·			-

AECOM DATA VALIDATION REPORT FORM – HEXAVALENT CHROMIUM ANALYSIS 7196 Page 5 of 5

Sample Calculations

	[156-MW7-201	60510] (JC19	912-1) pgs. 9, 130	
Background reading	0.00	14		
Total absorbance	0.00	14		
Total absorbance - background		0		
Instrument Response	0.00	10		
Sample volume (mL)	5	50		
Final Volume (mL)	5	50		
Percent solids		1		
Dilution Factor		1		
AECOM Calculated Result (mg/L)	0.0039U	OK	Reported Result (mg/L)	0.0039U

Client Name: PPG Industries			Project Number: 60493065			
Site Location: PPG Site 156 - Metro Towers, Jersey City, NJ			Project Manager: Nan Conrey			
Laboratory: SGS Accutest, Dayton, NJ			Type of Validation: Full			
Laboratory Job No: JC19972			Date Checked: 6/27/2016			
Validator: Dawn Brule			Peer: Mary Kozik			
ITEM YES NO		NC	N/A	COMMENTS		
Sample results included?	Х					
Paparting Limits mot project requirements?						

Reporting Limits met project requirements? Χ Field I.D. included? Χ Laboratory I.D. included? Did data package sample IDs match sample IDs Х on COC? Did electronic data deliverable (EDD) sample IDs Χ match COC sample IDs? Did data package sample IDs match electronic Χ data deliverable (EDD) sample IDs? Sample matrix included? Χ Χ Sample receipt temperature 2-6°C? 2.7°C Signed COCs included? Χ Date of sample collection included? Χ Date of sample digestion included? Χ Χ Date of analysis included? Holding time met QC criteria? (Metals -180 days from sample collection; Mercury - 28 days from Χ sample collection. If HT exceeded by 10 days R all results. Method reference included? Х Laboratory Case Narrative included? Χ

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?	х			156-MW6-20160510 (2x for TI)
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) -Blank plus 1 standard? If no, reject (R) data.			х	
3) Hg (7470/7471) -Blank plus 5 standards? If no, reject (R) data.			х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			x	N/A for Limited Validation
Analyzed immediately after initial calibration? If no, reject (R) data.			Х	
2) %R criteria met? (90-110%). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89%, and R all data for affected analyte(s) if %R <80% or >120%.			x	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
Analyzed immediately after each ICV/ICC/CB and after every 10 samples? If no, reject (R) data.			х	
2) CCS and CCV from independent source and at mid-level of calibration curve. If no, reject (R) data.			х	
3) %R criteria met? (90-110%R). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89% and R all data for affected analyte(s) if %R <80% or >120%.			x	
4) Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, Tl by 6010; 70-130% all others. If no, refer to ILM05.4 NJ SOP 5.A.2 for actions.			х	
Calibration Blanks			х	N/A for Limited Validation
Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples? If no, reject (R) data.			x	

ITEM	YES	NO	N/A	COMMENTS
2) Absolute value <3xIDL? If no, -if sample result <10x CB result, qualify affected analyte(s) in associated samples with CB; -if sample result >10xCB result, no qualification.			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? If no, reject (R) data, except no aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples? If - MB 1/25, J sample results from 21-25; -MB >1/25, R sample results after 25th sample.	х			
3) MB results nondetect? If no, -sample result <3xMB, negate UB; -sample result>3xMB but <10xMB, JB; -sample result >10xMB, no qualification.		х		<mdl< td=""></mdl<>
4) Negative MB result reported? If yes, -Positive sample result<10xMB, qualify estimated, biased low (J); -Non-detect sample result, qualify UJ, may be false non-detect.		х		<mdl< td=""></mdl<>
Field Blanks/Equipment Blanks Included in Lab Package?	х			156-FB20160510
1) FB/EB result non-detect? If no, -sample result <3xFB/EB, negate U; -sample result>3xFB/EB but <10xMB, J; -sample result >10xFB/EB, no qualification.		х		Ni=1.7B ug/l; V=0.80B ug/l
ICP Interference Check Sample (ICS) included in Lab Package?			х	N/A for Limited Validation
1) Analyzed at beginning of analytical run? If no, reject (R) data.			х	
2) %R criteria met? (80-120%) If no, %R>120%, no qualification if sample result non-detect; %R between 121-150%, J positive results, biased high; %R between 50-79%, J/UJ results, biased low; %R<50% or >150%, reject (R) result			х	
3) Spot check accuracy of %Rs			х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	х			MS not performed on site sample; batch QC not assessed
1) MS/MSD %R (75-125%R) and RPD (+20%) criteria met? - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; -%R<75% J/UJ for affected analyte(s) for all samples in the same batch/SDG; - RPD outside +20% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
3) Was the MS performed on a site sample?			х	
4) Was the MS performed on a FB/EB or TB? If yes, J all sample data.			х	
Post Digestion Spike			Х	
1) %R criteria met? (75-125%R) - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; - %R<75% J/UJ affected analyte(s) for all samples in the same batch/SDG.			х	
2) Was the spike performed on a FB/EB or TB? If yes, J all sample data.			х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	
Laboratory Duplicate Data Included in Lab Package?		Х		
Aqueous - If RPD is >20% but <100% and sample and duplicate results are >5x the QL, estimate (J) results >the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			x	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	х			
1) LCS %R criteria met? (80-120%R). If no, J/UJ all affected analytes(s) for all samples in the same batch/SDG.	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples? If no, J/UJ affected analyte(s) for all samples in the same batch/SDG.	х			
Serial Dilution	х			Not performed on site sample; batch QC not assessed
1) %D (<10%R) criteria met? - If analyte concentration >25xIDL (7000) or >10xIDL (6010) and %D >10% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was the frequency 1/batch or 20 samples?			Х	

ITEM	YES	NO	N/A	COMMENTS
3) Was a site sample used?			Х	
4) Was a FB/EB or TB used? If yes, J all sample data.			Х	
5) Spot check accuracy of %Ds.			Х	
Field Duplicate Data included in Lab Package?	Х			156-MW6-20160510 & 156-MW6D-20160510
Aqueous - If RPD is >20% but <100% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.		x		
Soil - If RPD is >35% but <120% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			х	
Percent Solids data included in Lab Package?			х	
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			х	

AECOM Page 6 of 6

Blanks

Equipment Blank	(ug/L)	(ug/L)	(ug/L)		
				<3X EB Negate (U) at RL, <10X EB Estimate (J), or >10x	
NICKEL	1.7	5.1	17	EB (OK)	All samples
				<3X EB Negate (UB), <10X EB Estimate (J), or >10x EB	All Samples
VANADIUM	0.80	2.4	8	(OK)	

Field Duplicate

					Dup	Dup			
Analyte	Sample ID	Result	RL	Dup Lab Sample ID	Result	RL	Unit	RPD	Action
ANTIMONY	156-MW6-20160510	3.3U	6	156-MW6D-0160510	3.3U	6	ug/l	NC	Both ND
CHROMIUM	156-MW6-20160510	1.7B	10	156-MW6D-0160510	1.7B	10	ug/l	0	<20%; ACCEPT
NICKEL	156-MW6-20160510	18.0	10	156-MW6D-0160510	16.6	10	ug/l	8.1	<20%; ACCEPT
THALLIUM	156-MW6-20160510	3.8U	4	156-MW6D-0160510	1.9U	2	ug/l	NC	Both ND
VANADIUM	156-MW6-20160510	17.7B	50	156-MW6D-0160510	16.3B	50	ug/l	8.2	<20%; ACCEPT



Data Validation Report

Project: PPG Site 156 - Metro Towers					
Laboratory:	SGS Accutest, D	ayton, NJ			
Laboratory Job No.:	.: JC20458				
Analysis/Method: Hexavalent Chromium SW846 3060A/7196 Metals by ICP-AES/ SW-846 6010C					
Validation Level:	Full (Hexavalent Limited (Metals)	Chromium)			
Site Location/Address	: PPG Site 156 - M	Metro Towers, Jersey City, NJ			
AECOM Project No:	60493065.20				
Prepared by: Dawn Brule /AECOM		Completed on: 07/07/2016			
Reviewed by: Mary Introduction	Kozik/AECOM	File Name: JC20458_2016-07-07_DV Report-F			

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP and/or Region 2 validation Standard Operating Procedure(s) (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, and USEPA SW-846 Method 7196A;
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods).
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015)

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.
- JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in

AECOM 2

the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.

RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on May 17, 2016 as part of the sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
FB-051716 (Equipment Blank)	JC20458-2	Aqueous	Hexavalent Chromium and Metals (Sb, Cr, Ni, V, Tl)
MW-04_051716	JC20458-1	IGround Water	Hexavalent Chromium and Metals (Sb, Cr, Ni, V, Tl)

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

MS Results

Sample MW-04_051716 was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The MS recovery was -2%; which did not meet the quality control criteria of 85-115% and was less than 50%; therefore, the hexavalent chromium results for the samples in this SDG were rejected on the basis of MS %Rs.

Metals

Equipment Blank

Nickel was detected in the equipment blank associated with the samples in this SDG, at a concentration above the MDL, but below the RL.

Nickel was detected at less than three times the EB in sample MW-04_051716; therefore, the result was negated at the RL.

AECOM 3

Sample Results

Reported results (flagged B by the laboratory) that were less than the RL, but greater than or equal to the MDL, are approximate values and have been qualified as estimated (J).

Data Quality and Usability

The results for hexavalent chromium in both samples were rejected; however, the results may be usable for project objectives as discussed below. Qualified results, if applicable, are presented in Attachments A and B.

Based on the MS recovery, the hexavalent chromium results in both samples in this SDG were rejected (RA). However, based on the reducing potential of the sample matrix shown by the Eh/pH phase diagram there is evidence to suggest that the matrix for this sample was reducing and not capable of supporting hexavalent chromium. Therefore, even though the sample results were rejected based on MS %Rs, these results may be usable for site decisions as estimated values.

Sample results considered to be negated due to blank contamination are usable as nondetect results at the RL.

Sample results reported between the MDL and RL are usable as estimated values.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hit List(s)

AECOM Page 1 of 2

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers

Sampling Date May 17, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC20458 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID FB-051716

Field Sample ID	Lab Sample ID	Analyte	Method Blank (mg/l)	,	Validation Sample Result (mg/l)	RL (mg/l)	Assurance	NJDEP Validation Footnote
MW-04_051716	JC20458-1	CHROMIUM (HEXAVALENT)	U	0.017	0.017	0.010	Reject	1
FB-051716	JC20458-2	CHROMIUM (HEXAVALENT)	U	U	U	0.010	Reject	1

Note: A "U" under Method Blank column indicates a nondetect result.

A "U" under the Laboratory Sample Result and Validation Sample Result columns indicates a nondetect result at the RL.

NJDEP Laboratory Footnote

1. The sample result was rejected because the matrix spike recovery was less than 50%.

AECOM Page 2 of 2

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 - Metro Towers

Sampling Date May 17, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC20458 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID FB-051716

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Laboratory Sample Result (ug/l)	Validation Sample Result (ug/l)	RL (ug/l)	Quality Assurance Decision	NJDEP Validation Footnote
MW-04_051716	JC20458-1	CHROMIUM	U	1.9B	1.9	10	Qualify	2
MW-04_051716	JC20458-1	NICKEL	U	2.5B	U	10	Negate	1
FB-051716	JC20458-2	NICKEL	U	5.1B	5.1	10	Qualify	2
MW-04_051716	JC20458-1	VANADIUM	U	2.4B	2.4	50	Qualify	2

Note: A "U" under Method Blank column indicates a nondetect result.

A "U" under the Laboratory Sample Result and Validation Sample Result columns indicates a nondetect result at the RL.

NJDEP Laboratory Footnote

- 1. The value reported is less than or equal to 3x the value in the trip/field blank. It is the policy of NJDEP-DPFSR to negate the reported value as due to probable foreign contamination unrelated to the actual sample. The end-user, however, is alerted that a reportable quantity of the analyte was detected.
- 2. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

Attachment B

Data Validation Report Form

Client Name: PPG Industries	Project Number: 60493065.20
Site Location: PPG Site 156 - Metro Towers, Jersey City, NJ	Project Manager: Nan Conrey
Laboratory: SGS Accutest, Dayton, NJ	Type of Validation: Full
Laboratory Job No: JC20458	Date Checked: 7/7/2016
Validator: Dawn Brule	Peer: Mary Kozik

ITEM	YES	NO	N/A	COMMENTS
Sample results included?	Х			
Reporting Limits met project requirements?	Х			
Field I.D. included?	Х			
Laboratory I.D. included?	Х			
Sample matrix included?	Х			
Sample receipt temperature 2-6°C?	Х			4.6°C
Signed COCs included?	Х			
Date of sample collection included?	Х			
Date of sample digestion included?	Х			
Holding time to digestion met criteria? (Soils -30 days from collection to digestion.)	Х			
Date of analysis included?	Х			
Holding time to analysis met criteria? (Soils -168 hours from digestion to analysis; Aqueous - 24 hours from collection to analysis.	х			
Method reference included?	Х			
Laboratory Case Narrative included?	Х			

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	х			
2) Correlation coefficient of =0.995 (7196A) or =0.999 (7199)	Х			
3) Calibrate daily or each time instrument is set up.	Х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	х			
1) %R criteria met? (90 - 110%)	Х			
2) Correct frequency of one per every 10 samples	х			
CCS and QCS from independent source and at mid-level of calibration curve	Х			
Calibration Blanks	х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	x			
2) Absolute value should not exceed MDL.	Х			
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	х			
Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	х			
1) Eh and pH data was included and plotted for all samples?	Х			
Soluble Matrix Spike Data Included in Lab Package?	Х			
1) Soluble Matrix %R criteria met? (75-125%R).		Х		See table below.
Was the spike concentration 40 mg/Kg or twice the sample concentration?		x		Spiked at 0.15 mg/L
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	х			
Insoluble Matrix Spike Data Included in Lab Package?			х	NA for aqueous samples
1) Insoluble Matrix %R criteria met? (75-125%R).			х	
2) Was the spike concentration around 400 to 800 mg/Kg?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			×	

Matrix Spikes

Sample ID	Analyte			Upper Limit	Action
MW-04_051716	CHROMIUM (HEXAVALENT)	-2	85	115	Reject (RA)

ODO# 1000450			⊐	
SDG#: JC20458	x - concentration	y - response		
Batch: GN46123	0	0.000		
Cr+6 ICAL 05/18/2016	0	0.002 0.012		
Aqueous	0.01	0.012		
(p. 122 of data pkg)	0.05	0.047		
	0.3	0.093		
	0.5	0.459		
	0.8	0.735		
	1	0.928		
		0.020		(p. 122 of data
AECOM Calculated Intercept	-0.0001	OK	Reported intercept	pkg) -0.0001
AECOM Slope	0.9226	OK	Reported Slope	0.9226
AECOM Calculated r	0.99994	OK	Reported r	0.99994
LCS calculation	GN46123-B1 pgs	. 118,122		
Background Absorbance	0			
Total absorbance	0.135			
Total absorbance - background	0.135			
Instrument Concentration	0.146			
Sample volume (mL)	50			
Final Volume (mL)	50			
Dilution Factor	1			
AECOM Calculated LCS Result (mg/L)	0.15	OK	Reported Result (mg/L)	0.15
%R = Found/True*100	pgs. 118,122			
True Value (mg/L)	0.15			
AECOM Calculated %R	97.6	OK, rounding	Reported %R	100.0
MS calculation	GN46123-S1 (JC2	20458-1) pgs. 120	0, 122	
Background reading	0.015			
Total absorbance	0.028			
Total absorbance - background	0.013			
Instrument Concentration	0.0142			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1			
AECOM Calculated MS Result (mg/L)	0.014	OK	Reported Result (mg/L)	0.014
%R = Found/True*100	GN46123-S1 (JC2	20458-1) pas. 120	0, 122	
True Value (mg/L)	0.15	, , ,		
Native concentration (mg/L)	0.017			
AECOM%R	-2	ОК	Reported %R	-2
Dan antino Limit	[BANA) 0.4 05.47403	(1000450.4)	- 0 400	
Reporting Limit Low Standard	[MW-04_051716]	(JC2U458-1) pg	S. J, 122	
COM STATICATO				
	0.01			
Initial volume (mL)	50			
Initial volume (mL) Final volume (mL)	50 50			
Initial volume (mL) Final volume (mL) Percent solids	50 50 1			
Initial volume (mL) Final volume (mL)	50 50	ОК	Reported RL (mg/L)=	0.010

AECOM DATA VALIDATION REPORT FORM – HEXAVALENT CHROMIUM ANALYSIS 7196 Page 6 of 6

Sample Calculations

	[MW-04_051716]	(JC2045	8-1) pgs. 9, 122	
Background reading	0.001			
Total absorbance	0.017			
Total absorbance - background	0.016			
Instrument Response	0.017			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1			
AECOM Calculated Result (mg/L)	0.017	OK	Reported Result (mg/L)	0.017

Client Name: PPG Industries	Project Number: 60493065.20		
Site Location: PPG Site 156 - Metro Towers, Jersey City, NJ	Project Manager: Nan Conrey		
Laboratory: SGS Accutest, Dayton, NJ	Type of Validation: Limited		
Laboratory Job No: JC20458	Date Checked: 7/7/2016		
Validator: Dawn Brule	Peer: Mary Kozik		

ITEM	YES	NO	N/A	COMMENTS
Sample results included?	Х			
Reporting Limits met project requirements?	Х			
Field I.D. included?	Х			
Laboratory I.D. included?	Х			
Did data package sample IDs match sample IDs on COC?	Х			
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	Х			
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	Х			
Sample matrix included?	Х			
Sample receipt temperature 2-6°C?	Х			4.6°C
Signed COCs included?	Х			
Date of sample collection included?	Х			
Date of sample digestion included?	Х			
Date of analysis included?	Х			
Holding time met QC criteria? (Metals -180 days from sample collection; Mercury - 28 days from sample collection. If HT exceeded by 10 days R all results.	x			
Method reference included?	Х			
Laboratory Case Narrative included?	Х			

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?		х		
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) -Blank plus 1 standard? If no, reject (R) data.			х	
3) Hg (7470/7471) -Blank plus 5 standards? If no, reject (R) data.			х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			x	N/A for Limited Validation
Analyzed immediately after initial calibration? If no, reject (R) data.			х	
2) %R criteria met? (90-110%). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89%, and R all data for affected analyte(s) if %R <80% or >120%.			х	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
Analyzed immediately after each ICV/ICC/CB and after every 10 samples? If no, reject (R) data.			х	
2) CCS and CCV from independent source and at mid-level of calibration curve. If no, reject (R) data.			х	
3) %R criteria met? (90-110%R). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89% and R all data for affected analyte(s) if %R <80% or >120%.			х	
Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, Tl by 6010; 70-130% all others. If no, refer to ILM05.4 NJ SOP 5.A.2 for actions.			х	
Calibration Blanks			х	N/A for Limited Validation
Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples? If no, reject (R) data.			х	

ITEM	YES	NO	N/A	COMMENTS
Absolute value <3xIDL? If no, -if sample result <10x CB result, qualify affected analyte(s) in associated samples with CB; -if sample result >10xCB result, no qualification.			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? If no, reject (R) data, except no aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples? If - MB 1/25, J sample results from 21-25; -MB >1/25, R sample results after 25th sample.	х			
3) MB results nondetect? If no, -sample result <3xMB, negate UB; -sample result>3xMB but <10xMB, JB; -sample result >10xMB, no qualification.		x		<mdl; no="" quals.<="" td=""></mdl;>
4) Negative MB result reported? If yes, -Positive sample result<10xMB, qualify estimated, biased low (J); -Non-detect sample result, qualify UJ, may be false non-detect.	х			<-mdl; no quals.
Field Blanks/Equipment Blanks Included in Lab Package?	х			FB-051716
The state of		х		Ni=5.1B ug/l; negate Ni result for samp MW-04_051716
ICP Interference Check Sample (ICS) included in Lab Package?			х	N/A for Limited Validation
Analyzed at beginning of analytical run? If no, reject (R) data.			х	
2) %R criteria met? (80-120%) If no, %R>120%, no qualification if sample result non-detect; %R between 121-150%, J positive results, biased high; %R between 50-79%, J/UJ results, biased low; %R<50% or >150%, reject (R) result			х	
3) Spot check accuracy of %Rs			х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	х			MS not performed on site sample; batch QC not assessed
1) MS/MSD %R (75-125%R) and RPD (+20%) criteria met? - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; -%R<75% J/UJ for affected analyte(s) for all samples in the same batch/SDG; - RPD outside +20% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
Was a sample spiked at the frequency of 1/batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
3) Was the MS performed on a site sample?			х	
4) Was the MS performed on a FB/EB or TB? If yes, J all sample data.			х	
Post Digestion Spike			х	
1) %R criteria met? (75-125%R) - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; - %R<75% J/UJ affected analyte(s) for all samples in the same batch/SDG.			х	
2) Was the spike performed on a FB/EB or TB? If yes, J all sample data.			х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	
Laboratory Duplicate Data Included in Lab Package?		х		
Aqueous - If RPD is >20% but <100% and sample and duplicate results are >5x the QL, estimate (J) results >the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			х	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	х			
1) LCS %R criteria met? (80-120%R). If no, J/UJ all affected analytes(s) for all samples in the same batch/SDG.	х			
Was an LCS analyzed at the frequency of 1/batch or 20 samples? If no, J/UJ affected analyte(s) for all samples in the same batch/SDG.	х			
Serial Dilution	х			Not performed on site sample; batch QC not assessed
1) %D (<10%R) criteria met? - If analyte concentration >25xIDL (7000) or >10xIDL (6010) and %D >10% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was the frequency 1/batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
3) Was a site sample used?			х	
4) Was a FB/EB or TB used? If yes, J all sample data.			Х	
5) Spot check accuracy of %Ds.			Х	
Field Duplicate Data included in Lab Package?		Х		
Aqueous - If RPD is >20% but <100% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			х	
Percent Solids data included in Lab Package?			х	
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			Х	

AECOM Page 6 of 6

Equipment Blank

Analyte	(ug/L)	3X EB (ug/L)	10X EB (ug/L)	Action	Associated samples
NICKEL	5.1	15.3	51	<3X EB Negate (U) at RL, <10X EB Estimate (J), or >10x EB (OK)	MW-04_051716



Data Validation Report

Project:	PPG Site 156 - M	etro Towers	
Laboratory:	SGS Accutest, Dayton, NJ		
Laboratory Job No.:	tory Job No.: JC22339		
Analysis/Method:	nium SW846 3060A/7196 S/ SW-846 6010C		
Validation Level: Full (Hexavalent Chrom Limited (Metals)		Chromium)	
Site Location/Address:	PPG Site 156 - M	etro Towers, Jersey City, NJ	
AECOM Project No:	60493065.20		
Prepared by: Dawn Brule /AECOM		Completed on: 07/07/2016; Revised 7/20/2016	
Reviewed by: Mary K	ozik/AECOM	File Name: JC22339_2016-07-07_DV Report-F	

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP and/or Region 2 validation Standard Operating Procedure(s) (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, and USEPA SW-846 Method 7196A:
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods).
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015)

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.

AECOM 2

JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.

RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on June 16, 2016 as part of the PPG Site 156 - Metro Towers sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
156-FB-20160616 (Equipment Blank)	JC22339-6	Aqueous	Hexavalent Chromium
156-FB-20160616 (Equipment Blank)	JC22339-6A	Aqueous	Metals (Sb, Cr, Ni, V, Tl)
156-MW1R-20160616	JC22339-4	Ground Water	Hexavalent Chromium
156-MW1R-20160616	JC22339-4A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW51R-20160616 (Field Duplicate of 156-MW1R-20160616)	JC22339-5	Ground Water	Hexavalent Chromium
156-MW51R-20160616 (Field Duplicate of 156-MW1R-20160616)	JC22339-5A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW5R-20160616	JC22339-3	Ground Water	Hexavalent Chromium
156-MW5R-20160616	JC22339-3A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW6-20160616	JC22339-1	Ground Water	Hexavalent Chromium
156-MW6-20160616	JC22339-1A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)
156-MW9-20160616	JC22339-2	Ground Water	Hexavalent Chromium
156-MW9-20160616	JC22339-2A	Ground Water	Metals (Sb, Cr, Ni, V, Tl)

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

MS Results

Sample 156-MW1R-20160616 was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The MS recovery was 2.7%; which did not meet the quality control criteria of 85-115% and was less than 50%;

AECOM 3

therefore, the hexavalent chromium results for all the samples in this SDG were rejected (RA) on the basis of MS %Rs.

Metals

ICP Serial Dilution Results

Sample 156-MW1R-20160616 was selected for the serial dilution. Chromium and nickel each had %differences greater than 10% but because sample results were less than 10x the IDLs, no qualifications were needed. Vanadium had a % difference greater than 10% and a sample result greater than 10x the IDL therefore, all samples with positive results for vanadium were qualified (J) as estimated.

Sample Results

Reported results (flagged B by the laboratory) that were less than the RL, but greater than or equal to the MDL, are approximate values and have been qualified as estimated (J).

Data Quality and Usability

The results for hexavalent chromium in all the samples in this SDG were rejected; however, the results may be usable for project objectives as discussed below. Qualified results are presented in Attachments A and B.

Based on the MS recovery, the hexavalent chromium results in all the samples in this SDG were rejected. However, based on the reducing potential of the sample matrix shown by the Eh/pH phase diagram there is evidence to suggest that the matrix for this sample was reducing and not capable of supporting hexavalent chromium. Therefore, even though the sample results were rejected based on MS %Rs, these results may be usable for site decisions as estimated values.

Sample results qualified due to poor ICP serial dilution results are usable as estimated values with an unknown directional bias.

Sample results reported between the MDL and RL are usable as estimated values.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hit List(s)

AECOM Page 1 of 3

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers

Sampling Date June 16, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC22339 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20160616

Field Sample ID	Lab Sample ID	Analyte	Method Blank (mg/l)	Sample Result	Validation Sample Result (mg/l)	RL (mg/l)	Assurance	NJDEP Validation Footnote
156-MW1R-20160616	JC22339-4	CHROMIUM (HEXAVALENT)	U	0.014	0.014	0.010	Reject	1
156-MW51R-20160616	JC22339-5	CHROMIUM (HEXAVALENT)	U	0.014	0.014	0.010	Reject	1
156-MW5R-20160616	JC22339-3	CHROMIUM (HEXAVALENT)	U	U	U	0.010	Reject	1
156-MW6-20160616	JC22339-1	CHROMIUM (HEXAVALENT)	U	0.0054B	0.0054	0.010	Reject	1,2
156-MW9-20160616	JC22339-2	CHROMIUM (HEXAVALENT)	U	0.050	0.050	0.010	Reject	1
156-FB-20160616	JC22339-6	CHROMIUM (HEXAVALENT)	U	U	U	0.010	Reject	1

Note: A "U" under Method Blank column indicates a nondetect result.

A "U" under the Laboratory Sample Result and Validation Sample Result columns indicates a nondetect result at the RL.

NJDEP Laboratory Footnote

- 1. The sample result was rejected because the matrix spike recovery was less than 50%.
- 2. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

AECOM Page 2 of 3

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 - Metro Towers

Sampling Date June 16, 2016 Lab Name/ID Accutest, Dayton, NJ

SDG No JC22339 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20160616

Field Sample ID	Lab Sample ID	Analyte	Method Blank (mg/kg)	Laboratory Sample Result (mg/kg)	Validation Sample Result (mg/kg)	RL (mg/kg)	Quality Assurance Decision	NJDEP Validation Footnote
156-MW6-20160616	JC22339-1A	CHROMIUM	U	4.0B	4.0	10	Qualify	2
156-MW9-20160616	JC22339-2A	CHROMIUM	U	1.9B	1.9	10	Qualify	2
156-MW5R-20160616	JC22339-3A	CHROMIUM	U	4.8B	4.8	10	Qualify	2
156-MW1R-20160616	JC22339-4A	CHROMIUM	U	4.7B	4.7	10	Qualify	2
156-MW51R-20160616	JC22339-5A	CHROMIUM	U	4.1B	4.1	10	Qualify	2
156-MW6-20160616	JC22339-1A	NICKEL	U	32.5	32.5	10		
156-MW5R-20160616	JC22339-3A	NICKEL	U	2.0B	2.0	10	Qualify	2
156-MW51R-20160616	JC22339-5A	NICKEL	U	1.2B	1.2	10	Qualify	2
156-MW6-20160616	JC22339-1A	VANADIUM	U	18.3B	18.3	50	Qualify	1,2
156-MW9-20160616	JC22339-2A	VANADIUM	U	1.7B	1.7	50	Qualify	1,2
156-MW5R-20160616	JC22339-3A	VANADIUM	U	2.1B	2.1	50	Qualify	1,2
156-MW1R-20160616	JC22339-4A	VANADIUM	U	7.2B	7.2	50	Qualify	1,2
156-MW51R-20160616	JC22339-5A	VANADIUM	U	6.7B	6.7	50	Qualify	1,2

Note: A "U" under Method Blank column indicates a nondetect result.

AECOM Page 3 of 3

NJDEP Laboratory Footnote

1. The reported value was qualified because serial dilution analysis was not within QC limit of 10% D.

2. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory

Attachment B

Data Validation Report Form

Client Name: PPG Industries	Project Number: 60493065.20
Site Location: PPG Site 156 - Metro Towers, Jersey City, NJ	Project Manager: Nan Conrey
Laboratory: SGS Accutest, Dayton, NJ	Type of Validation: Full
Laboratory Job No: JC22339	Date Checked: 7 /7/2016
Validator: Dawn Brule	Peer: Mary Kozik

YES NO N/A **COMMENTS ITEM** Sample results included? Χ Reporting Limits met project requirements? Χ Sample listed as 156-MW-6-20160616 on COC, Field I.D. included? Χ but report and edd listed it as 156-MW6-20160616 which is consistent with IDs of the other samples Laboratory I.D. included? Χ Χ Sample matrix included? Χ 4.3°C Sample receipt temperature 2-6°C? Χ Signed COCs included? Χ Date of sample collection included? Χ Date of sample digestion included? Holding time to digestion met criteria? (Soils -30 Χ days from collection to digestion.) Date of analysis included? Χ Holding time to analysis met criteria? (Soils -168 hours from digestion to analysis; Aqueous - 24 Χ hours from collection to analysis. Method reference included? Χ Laboratory Case Narrative included? Χ

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	х			
2) Correlation coefficient of =0.995 (7196A) or =0.999 (7199)	х			
3) Calibrate daily or each time instrument is set up.	Х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	х			
1) %R criteria met? (90 - 110%)	Х			
2) Correct frequency of one per every 10 samples	Х			
3) CCS and QCS from independent source and at mid-level of calibration curve	х			
Calibration Blanks	Х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	х			
2) Absolute value should not exceed MDL.	Х			
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	х			
Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	Х			
1) Eh and pH data was included and plotted for all samples?	х			
Soluble Matrix Spike Data Included in Lab Package?	х			
1) Soluble Matrix %R criteria met? (85-115%R).		х		See table below.
Was the spike concentration 40 mg/Kg or twice the sample concentration?		x		Spiked at 0.15 mg/L
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	х			
Insoluble Matrix Spike Data Included in Lab Package?			х	NA for aqueous samples
1) Insoluble Matrix %R criteria met? (75-125%R).			х	
2) Was the spike concentration around 400 to 800 mg/Kg?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			x	

Matrix Spikes

Sample ID	Analyte	% Recovery	Lower Limit	Upper Limit	Action
156-MW1R-20160616	CHROMIUM (HEXAVALENT)	2.7	85	115	Reject (RA)

SDG#: JC22339	x - concentration	y - response	7	
Batch: GN47627				
Cr+6 ICAL 06/16/2016	0	0		
Aqueous	0.01	0.009		
(p. 154 of data pkg)	0.05	0.047		
	0.1	0.092		
	0.3	0.27		
	0.5	0.469		
	0.8	0.743		
	1	0.918		
			_	(p. 154 of data pkg)
AECOM Calculated Intercept	0.0000	ОК	Reported intercept	0.0000
AECOM Slope	0.9232	OK	Reported Slope	0.9232
AECOM Calculated r	0.99991	OK	Reported r	0.99991
LCS calculation	GN47627-B2 pgs	. 150,154		
Background Absorbance	0	,		
Total absorbance	0.137			
Total absorbance - background	0.137			
Instrument Concentration	0.148			
Sample volume (mL)	50			
Final Volume (mL)	50			
Dilution Factor	1			
AECOM Calculated LCS Result (mg/L)	0.15	OK	Reported Result (mg/L)	0.15
	****		(3-)	****
%R = Found/True*100	pgs. 150,154			
True Value (mg/L)	0.15			
AECOM Calculated %R	98.9	OK, rounding	Reported %R	100.0
MS calculation	GN47627-S3 (JC2	22339-4) pgs. 152	2. 154	
Background reading	0.001	, , ,	•	
Total absorbance	0.018			
Total absorbance - background	0.017			
Instrument Concentration	0.0184			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1			
AECOM Calculated MS Result (mg/L)	0.018	OK	Reported Result (mg/L)	0.018
%R = Found/True*100	GN47627-53 (IC	2220-4) ngs 153	154	
True Value (mg/L)	GN47627-S3 (JC2 0.15		., IJT	
Native concentration (mg/L)	0.014			
AECOM%R	2.9	OK rounding	Reported %R	2.7
ALGONIAN	2.0	Orchounting	Reported 7010	2.1
Reporting Limit	[156-MW1R-2016	0616] (JC22339-4	1) pgs. 16, 154	
Low Standard	0.01			
Initial volume (mL)	50			
Final volume (mL)	50			
Percent solids				
Percent solids	1			
Dilution Factor	1 1			
		ОК	Reported RL (mg/L)=	0.010

AECOM DATA VALIDATION REPORT FORM – HEXAVALENT CHROMIUM ANALYSIS 7196 Page 6 of 6

Sample Calculations

	[156-MW1R-2016	0616] (.	JC22339-4) pgs. 16, 154	
Background reading	0.002			
Total absorbance	0.015			
Total absorbance - background	0.013			
Instrument Response	0.014			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1_			
AECOM Calculated Result (mg/L)	0.014	OK	Reported Result (mg/L)	0.014

Client Name: PPG Industries	Project Number: 60493065
Site Location: PPG Site 156 - Metro Towers, Jersey City, NJ	Project Manager: Nan Conrey
Laboratory: SGS Accutest, Dayton, NJ	Type of Validation: Limited
Laboratory Job No: JC22339	Date Checked: 7/7/2016
Validator: Dawn Brule	Peer: Mary Kozik

			•
YES	NO	N/A	COMMENTS
Х			
Х			
Х			
Х			
	х		Sample listed as 156-MW-6-20160616 on COC,
	х		but report and edd listed it as 156-MW6-20160616 which is consistent with IDs of the other samples
х			
Х			
Х			2.7°C
Х			
Х			
Х			
Х			
x			
Х			
Х			
	x x x x x x x x x x x x x	X X	X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X X

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?		х		
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) -Blank plus 1 standard? If no, reject (R) data.			х	
3) Hg (7470/7471) -Blank plus 5 standards? If no, reject (R) data.			Х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			x	N/A for Limited Validation
Analyzed immediately after initial calibration? If no, reject (R) data.			х	
2) %R criteria met? (90-110%). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89%, and R all data for affected analyte(s) if %R <80% or >120%.			х	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
Analyzed immediately after each ICV/ICC/CB and after every 10 samples? If no, reject (R) data.			х	
2) CCS and CCV from independent source and at mid-level of calibration curve. If no, reject (R) data.			х	
3) %R criteria met? (90-110%R). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89% and R all data for affected analyte(s) if %R <80% or >120%.			х	
Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, Tl by 6010; 70-130% all others. If no, refer to ILM05.4 NJ SOP 5.A.2 for actions.			х	
Calibration Blanks			х	N/A for Limited Validation
Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples? If no, reject (R) data.			х	

ITEM	YES	NO	N/A	COMMENTS
2) Absolute value <3xIDL? If no, -if sample result <10x CB result, qualify affected analyte(s) in associated samples with CB; -if sample result >10xCB result, no qualification.			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? If no, reject (R) data, except no aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples? If - MB 1/25, J sample results from 21-25; -MB >1/25, R sample results after 25th sample.	х			
3) MB results nondetect? If no, -sample result <3xMB, negate UB; -sample result>3xMB but <10xMB, JB; -sample result >10xMB, no qualification.		х		<mdl; no="" quals.<="" td=""></mdl;>
4) Negative MB result reported? If yes, -Positive sample result<10xMB, qualify estimated, biased low (J); -Non-detect sample result, qualify UJ, may be false non-detect.		х		no quals
Field Blanks/Equipment Blanks Included in Lab Package?	х			156-FB-20160616
1) FB/EB result non-detect? If no, -sample result <3xFB/EB, negate U; -sample result>3xFB/EB but <10xMB, J; -sample result >10xFB/EB, no qualification.	х			
ICP Interference Check Sample (ICS) included in Lab Package?			х	N/A for Limited Validation
Analyzed at beginning of analytical run? If no, reject (R) data.			х	
2) %R criteria met? (80-120%) If no, %R>120%, no qualification if sample result non-detect; %R between 121-150%, J positive results, biased high; %R between 50-79%, J/UJ results, biased low; %R<50% or >150%, reject (R) result			Х	
3) Spot check accuracy of %Rs			х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	х			JC22339-4A
1) MS/MSD %R (75-125%R) and RPD (+20%) criteria met? - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; -%R<75% J/UJ for affected analyte(s) for all samples in the same batch/SDG; - RPD outside +20% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.	х			
2) Was a sample spiked at the frequency of 1/batch or 20 samples?	х			

ITEM	YES	NO	N/A	COMMENTS
3) Was the MS performed on a site sample?	х			
4) Was the MS performed on a FB/EB or TB? If yes, J all sample data.		Х		
Post Digestion Spike			х	
1) %R criteria met? (75-125%R) - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; - %R<75% J/UJ affected analyte(s) for all samples in the same batch/SDG.			Х	
2) Was the spike performed on a FB/EB or TB? If yes, J all sample data.			х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	
Laboratory Duplicate Data Included in Lab Package?		Х		
Aqueous - If RPD is >20% but <100% and sample and duplicate results are >5x the QL, estimate (J) results >the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			х	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	Х			
1) LCS %R criteria met? (80-120%R). If no, J/UJ all affected analytes(s) for all samples in the same batch/SDG.	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples? If no, J/UJ affected analyte(s) for all samples in the same batch/SDG.	Х			
Serial Dilution	х			JC22339-4A
1) %D (<10%R) criteria met? - If analyte concentration >25xIDL (7000) or >10xIDL (6010) and %D >10% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.		Х		V=15.3%; J positive results
2) Was the frequency 1/batch or 20 samples?	Х			

ITEM	YES	NO	N/A	COMMENTS
3) Was a site sample used?	Х			
4) Was a FB/EB or TB used? If yes, J all sample data.		Х		
5) Spot check accuracy of %Ds.			Х	
Field Duplicate Data included in Lab Package?	Х			JC22339-4A & JC22339-5A
Aqueous - If RPD is >20% but <100% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.	x			Ni: abs diff < RL; no qual.
Soil - If RPD is >35% but <120% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			х	
Percent Solids data included in Lab Package?			х	Aqueous samples
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			х	

AECOM Page 6 of 6

ICP Serial Dilution

JC22339-4A Analyte	Sample (ug/L)	Serial dilution (ug/L)	% Difference	10x IDL (ug/L)	Action
VANADIUM	7.20	8.30	15.3	5	J positive results



Data Validation Report

Project:		PPG Site 156 - Metro	Towers			
Laboratory:		SGS/Accutest, Dayton, NJ				
Laboratory Job	No.:	JC24082				
Analysis/Method: Hexavalent Chromium Metals by ICP-AES/ S'						
Validation Level: Full (Hexavalent Chror Limited (Metals)			nium)			
Site Location/A	Address:	PPG Site 156 - Metro	Towers , Jersey City, NJ			
AECOM Projec	ct No:	60493065.20				
Prepared by:		le/AECOM :herford/AECOM	Completed on: 07/26/2016			
Reviewed by:	Mary Kozi	k/AECOM	File Name: JC24082_2016-07-26_DV Report-FREV			
Introduction						

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP validation Standard Operating Procedures (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, and USEPA SW-846 Method 7196A;
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods);
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015).

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.

AECOM 2

JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

- R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.
- RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on July 14, 2016 as part of the sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
156-FB-20160714 (Equipment Blank)	JC24082-2	lAqueous	Hexavalent Chromium and Metals (Sb, Cr, Ni, Tl, V)
156-MW4-20160714	JC24082-1	Ground Water	Hexavalent Chromium and Metals (Sb, Cr, Ni, Tl, V)

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit List(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Boron, instead of vanadium, was mistakenly requested on the chain of custody (COC). The results for vanadium were retrieved by the laboratory and submitted as a revised data report.

Hexavalent Chromium

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

MS Results

Sample 156-MW4-20160714 was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The MS recovery was 97.3%; which met the quality control criteria of 85-115%. No data qualification was required on the basis of spike recovery.

AECOM 3

<u>Metals</u>

Sample Results

Reported results (flagged B by the laboratory) that were less than the RL, but greater than or equal to the MDL, are approximate values and have been qualified as estimated (J).

Data Quality and Usability

In general, these data appear to be valid and may be used for decision-making purposes. No data were rejected. Qualified results, if applicable, are presented in Attachments A and B.

Sample results reported between the MDL and RL are usable as estimated values with an unknown directional bias.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hit List(s)

AECOM Page 1 of 2

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers

Sampling Date July 14, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC24082 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20160714

Field Sample ID	Lab Sample ID		Blank	Sample Result	Validation Sample Result (mg/l)	RL (mg/l)	Assurance	NJDEP Validation Footnote
156-MW4-20160714	JC24082-1	CHROMIUM (HEXAVALENT)	U	0.044	0.044	0.010		

Note: A "U" under Method Blank column indicates a nondetect result.

AECOM Page 2 of 2

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 - Metro Towers

Sampling Date July 14, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC24082 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20160714

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Sample Result	Validation Sample Result (ug/l)	RL (ug/l)	Assurance	NJDEP Validation Footnote
156-MW4-20160714	JC24082-1	CHROMIUM	U	1.8B	1.8	10	Qualify	1
156-MW4-20160714	JC24082-1	NICKEL	U	1.4B	1.4	10	Qualify	1
156-MW4-20160714	JC24082-1	VANADIUM	U	2.0B	2.0	50	Qualify	1

Note: A "U" under Method Blank column indicates a nondetect result.

NJDEP Laboratory Footnote

1. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

Attachment B

Data Validation Report Form

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation: Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	х			
2) Correlation coefficient of =0.995 (7196A) or =0.999 (7199)	х			
3) Calibrate daily or each time instrument is set up.	х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	х			
1) %R criteria met? (90 - 110%)	х			
2) Correct frequency of one per every 10 samples	х			
3) CCS and QCS from independent source and at mid-level of calibration curve	Х			
Calibration Blanks	х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	Х			
2) Absolute value should not exceed MDL.	х			
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	x			
Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	Х			
1) Eh and pH data was included and plotted for all samples?	Х			
Soluble Matrix Spike Data Included in Lab Package?	Х			JC24082-1
1) Soluble Matrix %R criteria met? (85-115%R).	Х			
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?		х		As an aqueous sample, the spike concentration was 0.15 mg/l. The data was not affected.
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	Х			
Insoluble Matrix Spike Data Included in Lab Package?			Х	NA – aqueous samples
1) Insoluble Matrix %R criteria met? (75-125%R).			х	
2) Was the spike concentration around 400 to 800 mg/Kg?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	

AECOM Page 4 of 5

SDG#: JC24082	x - concentration	y - response		
Batch: GN49024				
Cr+6 ICAL 07/14/2016	0	0		
Aqueous	0.01	0.011		
(p. 194 of data pkg)	0.05	0.054		
	0.1	0.1		
	0.3	0.248		
	0.5	0.469		
	0.8	0.741		
	1	0.909		(p. 194 of data
				pkg)
AECOM Calculated Intercept	0.0015	OK	Reported intercept	0.0015
AECOM Slope	0.9131	OK	Reported Slope	0.9131
AECOM Calculated r	0.99937	OK	Reported r	0.99937
LCS calculation	GN49024-B1 pgs	190 194		
Background Absorbance	0 Oly49024-D1	. 100,104		
Total absorbance	0.135			
Total absorbance - background	0.135			
Instrument Concentration	0.146			
Sample volume (mL)	50			
Final Volume (mL)	50			
Dilution Factor	1			
AECOM Calculated LCS Result (mg/L)	0.15	OK	Reported Result (mg/L)	0.15
%R = Found/True*100	pgs. 190,194			
True Value (mg/L)	0.15			7
AECOM Calculated %R	97.5	OK, rounding	Reported %R	100
MS calculation	GN49024-S1 (JC	24082-1) pgs. 19	2, 194	
Background reading	0	,	,	
Total absorbance	0.179			
Total absorbance - background	0.179			
Instrument Concentration	0.1944			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1			
AECOM Calculated MS Result (mg/L)	0.19	OK	Reported Result (mg/L)	0.19
%R = Found/True*100	GN49024-S1 (JC	24082-1) nas 19	2 194	
True Value (mg/L)	0.15		-, IVT	
Native concentration (mg/L)	0.044			
AECOM%R	100	OK rounding	Reported %R	97
		_		
Reporting Limit	[156-MW4-20160	714] (JC24082-1)	pgs. 8, 194	
Low Standard	0.01			
Initial volume (mL)	50			
Final volume (mL)	50			
Percent solids	1			
Dilution Factor	1_			
Reporting Limit (mg/L)	0.010	OK	Reported RL (mg/L)=	0.010

AECOM Page 5 of 5

0.044

Sample Calculations

AECOM Calculated Result (mg/L)

	[156-MW4-20160714] (JC24082-1) pgs. 8, 194	
Background reading	0	
Total absorbance	0.042	
Total absorbance - background	0.042	
Instrument Response	0.044	
Sample volume (mL)	50	
Final Volume (mL)	50	
Percent solids	1	
Dilution Factor	1	

OK

Reported Result (mg/L)

0.044

Date of analysis included?

Holding time met QC criteria?

Method reference included?

Laboratory Case Narrative included?

Client Name: PPG Industries			Project Number: 60493065.20				
Site Location: PPG Site 156 - Metro Towers				Project Manager: Nanette Conrey			
Laboratory: SGS/Accutest, Dayton, NJ				Type of Validation: Limited			
Laboratory Job No: JC24082			Date Checked: 07/26/2016				
Validator: Dawn Brule, Kristin Rutherford			Peer: Mary Kozik				
ITEM	YES N		O N/A	N/A COMMENTS			
Sample results included?				Note: Boron mistakenly requested on COC instead of Vanadium. Data revised to includ instead of B on 7/25/16.	e V		
Reporting Limits met project requirements?	Х						
Field I.D. included?	Х						
Laboratory I.D. included?	Х						
Did data package sample IDs match sample IDs on COC?							
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	х						
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	х						
Sample matrix included?	Х						
Sample receipt temperature 2-6°C?				3.9°C			
Signed COCs included?							
Date of sample collection included?							
Date of sample digestion included?							

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

Χ

Χ

Χ

Χ

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?		х		
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
1) Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) - plus 1 standard?			х	
3) Cn and Hg -Blank plus 5 standards?			х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			x	N/A for Limited Validation
1) Analyzed immediately after initial calibration?			х	
2) %R criteria met? (90-110%).			х	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
1) Analyzed immediately after each ICV/ICC/CB and after every 10 samples?			х	
2) CCS and CCV from independent source and at mid-level of calibration curve?			х	
3) %R criteria met? (90-110%R).			Х	
4) Spot check CCV/CCS results for several analytes.			Х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, TI by 6010; 70-130% all others.			x	
Calibration Blanks			х	N/A for Limited Validation
Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples?			х	
2) Absolute value <3xIDL?			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? No aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples?	х			
3) MB results nondetect?		х		<mdl; no="" quals.<="" td=""></mdl;>
4) Negative MB result reported?	Х			<mdl; no="" quals<="" td=""></mdl;>

ITEM	YES	NO	N/A	COMMENTS
Field Blanks/Equipment Blanks Included in Lab Package?	х			
1) FB/EB result non-detect?	х			
ICP Interference Check Sample (ICS) included in Lab Package?			х	N/A for Limited Validation
Analyzed at beginning of analytical run? If no, reject (R) data.			Х	
2) %R criteria met? (80-120%)			х	
3) Spot check accuracy of %Rs			х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	Х			JC24082-1
1) MS/MSD %R (75-125%R) and RPD (<20%) criteria met?	х			
2) Was a sample spiked at the frequency of 1/batch or 20 samples?	х			
3) Was the MS performed on a site sample?	х			
4) Was the MS performed on a FB/EB or TB?		х		
Post Digestion Spike			х	
1) %R criteria met? (75-125%R)			Х	
2) Was the spike performed on a FB/EB or TB? If yes, J all sample data.			x	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			Х	
Laboratory Duplicate Data Included in Lab Package?		х		
Aqueous – Is RPD ≤ 20%?			Х	
Soil – Is RPD ≤ 35%?			Х	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	x			
1) LCS %R criteria met? (80-120%R).	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	х			
Serial Dilution	х			JC24082-1
1) %D (<10%R) criteria met?		х		Cr, Ni, & V <10x IDL; no quals
2) Was the frequency 1/batch or 20 samples?	х			
3) Was a site sample used?	х			

ITEM	YES	NO	N/A	COMMENTS
4) Was a FB/EB or TB used?		Х		
5) Spot check accuracy of %Ds.			Х	
Field Duplicate Data included in Lab Package?		Х		
Aqueous - Is RPD is <20%?			х	
Soil - Is RPD is <35%?			х	
Percent Solids data included in Lab Package?			Х	Aqueous samples
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			Х	



Data Validation Report

Project:	PPG Site 156 - Metro Towers						
Laboratory:	SGS/Accutest, Dayton,	SGS/Accutest, Dayton, NJ					
Laboratory Job No.:	JC22453						
Analysis/Method:	Metals by ICP-AES/ SW846-6010 Hexavalent Chromium SW846 3060A/7196 Metals by ICP-AES/ SW846-6010						
Validation Level:	ium)						
Site Location/Address:	PPG Site 156 - Metro T	owers , Jersey City, NJ					
AECOM Project No:	60493065.20						
Prepared by: Dawn Brule /AECOM		Completed on: 07/12/2016					
Reviewed by: Mary Kozik/AECOM		File Name: JC22453_2016-07-12_DV Report-F					

Introduction

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP validation Standard Operating Procedures (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, and USEPA SW-846 Method 7196A:
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods);
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015)

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.

AECOM 2

JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.

RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on June 17, 2016 as part of the sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
156-MW10-20160617	JC22453-3	Ground Water	Hexavalent Chromium
156-MW10-20160617	JC22453-3A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)
156-MW3-20160617	JC22453-2	Ground Water	Hexavalent Chromium
156-MW3-20160617	JC22453-2A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)
156-MW7-20160617	JC22453-4	Ground Water	Hexavalent Chromium
156-MW7-20160617	JC22453-4A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)
156-MW8-20160617	JC22453-1	Ground Water	Hexavalent Chromium
156-MW8-20160617	JC22453-1A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

MS Results

Sample 156-MW10-20160617 was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The MS recovery was 113.3%; which met the quality control criteria of 85-115%. No data qualification was required on the basis of spike recovery.

AECOM 3

<u>Metals</u>

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

Data Quality and Usability

In general, these data appear to be valid and may be used for decision-making purposes. No data were rejected. Qualified results, if applicable, are presented in Attachments A and B.

Sample results reported between the MDL and RL are usable as estimated values with an unknown directional bias.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hitlist(s)

AECOM Page 1 of 2

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers

Sampling Date June 17, 2016

Lab Name/ID Accutest, Dayton, NJ

SDG No JC22453
Sample Matrix Aqueous
Trip Blank ID NA
Field Blank ID NA

Field Sample ID	Lab Sample ID	Analyte	Method Blank (mg/l)	Laboratory Sample Result (mg/l)	Validation Sample Result (mg/l)			NJDEP Validation Footnote
156-MW3-20160617	JC22453-2	CHROMIUM (HEXAVALENT)	U	0.0069B	0.0069	0.010	Qualify	1
156-MW8-20160617	JC22453-1	CHROMIUM (HEXAVALENT)	U	0.0058B	0.0058	0.010	Qualify	1

Note: A "U" under Method Blank column indicates a nondetect result.

NJDEP Laboratory Footnote

1. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

AECOM Page 2 of 2

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 - Metro Towers

Sampling Date June 17, 2016 Lab Name/ID Accutest, Dayton, NJ

SDG No JC22453
Sample Matrix Aqueous
Trip Blank ID NA
Field Blank ID NA

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Laboratory Sample Result (ug/l)	Validation Sample Result (ug/l)	RL (ug/l)	Quality Assurance Decision	NJDEP Validation Footnote
156-MW8-20160617	JC22453-1A	CHROMIUM	U	730	730	10		
156-MW3-20160617	JC22453-2A	CHROMIUM	U	2.7B	2.7	10	Qualify	1
156-MW10-20160617	JC22453-3A	CHROMIUM	U	5.3B	5.3	10	Qualify	1
156-MW7-20160617	JC22453-4A	CHROMIUM	U	2.0B	2.0	10	Qualify	1
156-MW8-20160617	JC22453-1A	NICKEL	U	3.7B	3.7	10	Qualify	1
156-MW3-20160617	JC22453-2A	NICKEL	U	1.3B	1.3	10	Qualify	1
156-MW10-20160617	JC22453-3A	NICKEL	U	62.7	62.7	10		
156-MW7-20160617	JC22453-4A	NICKEL	U	2.9B	2.9	10	Qualify	1
156-MW3-20160617	JC22453-2A	THALLIUM	U	1.9B	1.9	2.0	Qualify	1
156-MW8-20160617	JC22453-1A	VANADIUM	U	13.2B	13.2	50	Qualify	1
156-MW3-20160617	JC22453-2A	VANADIUM	U	3.4B	3.4	50	Qualify	1
156-MW10-20160617	JC22453-3A	VANADIUM	U	5.2B	5.2	50	Qualify	1
156-MW7-20160617	JC22453-4A	VANADIUM	U	2.1B	2.1	50	Qualify	1

Note: A "U" under Method Blank column indicates a nondetect result.

NJDEP Laboratory Footnote

1. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

Attachment B

Data Validation Report Form

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation: Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	х			
2) Correlation coefficient of =0.995 (7196A) or =0.999 (7199)	х			
Calibrate daily or each time instrument is set up.	х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	х			
1) %R criteria met? (90 - 110%)	Х			
2) Correct frequency of one per every 10 samples	х			
3) CCS and QCS from independent source and at mid-level of calibration curve	Х			
Calibration Blanks	х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	Х			
2) Absolute value should not exceed MDL.	х			
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	x			
Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	Х			
1) Eh and pH data was included and plotted for all samples?	Х			
Soluble Matrix Spike Data Included in Lab Package?	Х			JC22453-3
1) Soluble Matrix %R criteria met? (85-115%R).	Х			
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?		x		0.15 mg/l
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	Х			
Insoluble Matrix Spike Data Included in Lab Package?			Х	NA – aqueous samples
1) Insoluble Matrix %R criteria met? (75-125%R).			Х	
2) Was the spike concentration around 400 to 800 mg/Kg?			Х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	

AECOM Page 4 of 5

SDG#: JC22453	x - concentration	y - response		
Batch: GN47687				
Cr+6 ICAL 06/17/2016	0	0		
Aqueous	0.01	0.01		
(p. 118 of data pkg)	0.05	0.05		
	0.1	0.1		
	0.3	0.29		
	0.5	0.468		
	0.8	0.753		
	1	0.934	J	/n 110 of data
				(p. 118 of data pkg)
AECOM Calculated Intercept	0.0035	OK	Reported intercept	0.0035
AECOM Slope	0.9336	OK	Reported Slope	0.9336
AECOM Calculated r	0.99995	OK	Reported r	0.99995
LCS calculation	GN47687-B1 pgs	. 114,118		
Background Absorbance	0			
Total absorbance	0.15			
Total absorbance - background	0.15			
Instrument Concentration	0.157			
Sample volume (mL)	50			
Final Volume (mL)	50			
Dilution Factor	1			
AECOM Calculated LCS Result (mg/L)	0.16	OK	Reported Result (mg/L)	0.16
%R = Found/True*100	pgs. 114,118			
True Value (mg/L)	0.15			
AECOM Calculated %R	104.6	OK, rounding	Reported %R	106.7
MS calculation	GN47687-S3 (JC2	22453-3) pgs. 116	6, 118	
Background reading	0.006			
Total absorbance	0.167			
Total absorbance - background				
Instrument Concentration	0.161			
	0.161 0.1687			
Sample volume (mL)	0.1687 50			
Sample volume (mL) Final Volume (mL)	0.1687			
	0.1687 50			
Final Volume (mL) Percent solids Dilution Factor	0.1687 50 50 1 1			
Final Volume (mL) Percent solids	0.1687 50 50 1	ОК	Reported Result (mg/L)	0.17
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L)	0.1687 50 50 1 1 0.17			0.17
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100	0.1687 50 50 1 1 0.17			0.17
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L)	0.1687 50 50 1 1 0.17 GN47687-S3 (JC2 0.15			0.17
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L)	0.1687 50 50 1 1 0.17 GN47687-S3 (JC2 0.15	22453-3) pgs. 116	5, 118	
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L)	0.1687 50 50 1 1 0.17 GN47687-S3 (JC2 0.15			0.17
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L) AECOM%R Reporting Limit	0.1687 50 50 1 1 0.17 GN47687-S3 (JC 2 0.15 0 112.4	22453-3) pgs. 116	6, 118 Reported %R	
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L) AECOM%R Reporting Limit Low Standard	0.1687 50 50 1 1 0.17 GN47687-S3 (JC 2 0.15 0 112.4	22453-3) pgs. 116	6, 118 Reported %R	
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L) AECOM%R Reporting Limit Low Standard Initial volume (mL)	0.1687 50 50 1 1 0.17 GN47687-S3 (JC 2 0.15 0 112.4	22453-3) pgs. 116	6, 118 Reported %R	
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L) AECOM%R Reporting Limit Low Standard	0.1687 50 50 1 1 0.17 GN47687-S3 (JC 2 0.15 0 112.4	22453-3) pgs. 116	6, 118 Reported %R	
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L) AECOM%R Reporting Limit Low Standard Initial volume (mL)	0.1687 50 50 1 1 0.17 GN47687-S3 (JC: 0.15 0 112.4 [156-MW10-2016(0.01 50	22453-3) pgs. 116	6, 118 Reported %R	
Final Volume (mL) Percent solids Dilution Factor AECOM Calculated MS Result (mg/L) %R = Found/True*100 True Value (mg/L) Native concentration (mg/L) AECOM%R Reporting Limit Low Standard Initial volume (mL) Final volume (mL)	0.1687 50 50 1 1 0.17 GN47687-S3 (JC2 0.15 0 112.4 [156-MW10-20166 0.01 50 50	22453-3) pgs. 116	6, 118 Reported %R	

AECOM Page 5 of 5

Sample Calculations

	[156-MW10-20	160617] (JC22	453-3) pgs. 13, 118	
Background reading	0.00)6		
Total absorbance	0.00)8		
Total absorbance - background	0.00)2		
Instrument Response	-0.00)2		
Sample volume (mL)	5	50		
Final Volume (mL)	5	50		
Percent solids		1		
Dilution Factor		1		
AECOM Calculated Result (mg/L)	0.0039 U	OK	Reported Result (mg/L)	0.0039 U

Client Name: PPG Industries			Project Number: 60493065.20				
Site Location: PPG Site 156 - Metro Towers			Project Manager: Nanette Conrey				
Laboratory: SGS/Accutest, Dayton, NJ			Type of Validation: Limited				
Laboratory Job No: JC22453		0	Date C	checked: 07/12/2016			
Validator: Dawn Brule		F	Peer: N	Mary Kozik			
ITEM	YES	NO	N/A	COMMENTS			
Sample results included?	Х						
Reporting Limits met project requirements?	Х						
Field I.D. included?	Х						
Laboratory I.D. included?	Х						
Did data package sample IDs match sample IDs on COC?	Х						
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	х						
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	х						
Sample matrix included?	Х						
Sample receipt temperature 2-6°C?	Х			5°C			
Signed COCs included?	Х						
Date of sample collection included?	Х						
Date of sample digestion included?	Х						
Date of analysis included?	Х						
Holding time met QC criteria? (Metals -180 days from sample collection; Mercury - 28 days from sample collection. If HT exceeded by 10 days R all results.	x						
Method reference included?	Х						
Laboratory Case Narrative included?	Х						

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?		х		
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
1) Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) -Blank plus 1 standard? If no, reject (R) data.			х	
3) Hg (7470/7471) -Blank plus 5 standards? If no, reject (R) data.			х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			x	N/A for Limited Validation
1) Analyzed immediately after initial calibration? If no, reject (R) data.			x	
2) %R criteria met? (90-110%). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89%, and R all data for affected analyte(s) if %R <80% or >120%.			X	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
1) Analyzed immediately after each ICV/ICC/CB and after every 10 samples? If no, reject (R) data.			х	
2) CCS and CCV from independent source and at mid-level of calibration curve. If no, reject (R) data.			х	
3) %R criteria met? (90-110%R). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89% and R all data for affected analyte(s) if %R <80% or >120%.			х	
4) Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, Tl by 6010; 70-130% all others. If no, refer to ILM05.4 NJ SOP 5.A.2 for actions.			х	
Calibration Blanks			х	N/A for Limited Validation
1) Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples? If no, reject (R) data.			х	

ITEM	YES	NO	N/A	COMMENTS
2) Absolute value <3xIDL? If no, -if sample result <10x CB result, qualify affected analyte(s) in associated samples with CB; -if sample result >10xCB result, no qualification.			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? If no, reject (R) data, except no aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples? If - MB 1/25, J sample results from 21-25; -MB >1/25, R sample results after 25th sample.	х			
3) MB results nondetect? If no, -sample result <3xMB, negate UB; -sample result>3xMB but <10xMB, JB; -sample result >10xMB, no qualification.		х		<mdl; no="" quals.<="" td=""></mdl;>
4) Negative MB result reported? If yes, -Positive sample result<10xMB, qualify estimated, biased low (J); -Non-detect sample result, qualify UJ, may be false non-detect.	x			>-mdl; no quals
Field Blanks/Equipment Blanks Included in Lab Package?		х		
The state of			х	
ICP Interference Check Sample (ICS) included in Lab Package?			x	N/A for Limited Validation
1) Analyzed at beginning of analytical run? If no, reject (R) data.			х	
2) %R criteria met? (80-120%) If no, %R>120%, no qualification if sample result non-detect; %R between 121-150%, J positive results, biased high; %R between 50-79%, J/UJ results, biased low; %R<50% or >150%, reject (R) result			х	
3) Spot check accuracy of %Rs			х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	х			MS not performed on site sample; batch QC not assessed
1) MS/MSD %R (75-125%R) and RPD (+20%) criteria met? - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; -%R<75% J/UJ for affected analyte(s) for all samples in the same batch/SDG; - RPD outside +20% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	

ІТЕМ	YES	NO	N/A	COMMENTS
3) Was the MS performed on a site sample?			х	
4) Was the MS performed on a FB/EB or TB? If yes, J all sample data.			х	
Post Digestion Spike			х	
1) %R criteria met? (75-125%R) - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; - %R<75% J/UJ affected analyte(s) for all samples in the same batch/SDG.			х	
2) Was the spike performed on a FB/EB or TB? If yes, J all sample data.			х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	
Laboratory Duplicate Data Included in Lab Package?		х		
Aqueous - If RPD is >20% but <100% and sample and duplicate results are >5x the QL, estimate (J) results >the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			x	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	Х			
1) LCS %R criteria met? (80-120%R). If no, J/UJ all affected analytes(s) for all samples in the same batch/SDG.	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples? If no, J/UJ affected analyte(s) for all samples in the same batch/SDG.	х			
Serial Dilution	Х			Not performed on site sample; batch QC not assessed.
1) %D (<10%R) criteria met? - If analyte concentration >25xIDL (7000) or >10xIDL (6010) and %D >10% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was the frequency 1/batch or 20 samples?			Х	

ITEM	YES	NO	N/A	COMMENTS
3) Was a site sample used?			х	
4) Was a FB/EB or TB used? If yes, J all sample data.			х	
5) Spot check accuracy of %Ds.			х	
Field Duplicate Data included in Lab Package?		Х		
Aqueous - If RPD is >20% but <100% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			x	
Percent Solids data included in Lab Package?			х	Aqueous samples
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			х	



Data Validation Report

Project:	PPG Site 156 - Metro Towers					
Laboratory:	SGS/Accutest, Dayton, NJ					
Laboratory Job No.:	JC20060					
Analysis/Method: Hexavalent Chromium SW846 3060A/7196 Metals by ICP-AES/ SW846-6010						
Validation Level:	Full (Hexavalent Chrom Limited (Metals)	ium)				
Site Location/Address:	PPG Site 156 - Metro Te	owers , Jersey City, NJ				
AECOM Project No:	60493065.20					
Prepared by: Dawn Brule /AECOM		Completed on: 06/30/2016				
Reviewed by: Mary Kozik/AECOM		File Name: JC20060_2016-06-30_DV Report-F				

Introduction

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP validation Standard Operating Procedures (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, and USEPA SW-846 Method 7196A:
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods);
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015)

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.

AECOM 2

JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.

RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on May 11, 2016 as part of the sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
156-FB-20160511 (Equipment Blank)	JC20060-4	Aqueous	Hexavalent Chromium
156-FB-20160511 (Equipment Blank)	JC20060-4A	Aqueous	Metals (Sb, Cr, Ni, Tl, V)
156-MW10-20160511	JC20060-5	Ground Water	Hexavalent Chromium
156-MW10-20160511	JC20060-5A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)
156-MW3-20160511	JC20060-1	Ground Water	Hexavalent Chromium
156-MW3-20160511	JC20060-1A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)
156-MW8-20160511	JC20060-2	Ground Water	Hexavalent Chromium
156-MW8-20160511	JC20060-2A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)
156-MW9-20160511	JC20060-3	Ground Water	Hexavalent Chromium
156-MW9-20160511	JC20060-3A	Ground Water	Metals (Sb, Cr, Ni, Tl, V)

The samples were collected following the procedures detailed in the Work Order for Remedial Action at PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

AECOM 3

MS Results

Sample 156-MW10-20160511 was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The MS recovery was 86.7%, which met the quality control criteria of 85-115%. No data qualification was required on the basis of spike recoveries.

Metals

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

Data Quality and Usability

In general, these data appear to be valid and may be used for decision-making purposes. No data were rejected. Qualified results, if applicable, are presented in Attachments A and B.

Sample results reported between the MDL and RL are usable as estimated values with an unknown directional bias.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hitlist(s)

AECOM Page 1 of 2

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers

Sampling Date May 11, 2016

Lab Name/ID SGS/Accutest, Dayton, NJ

SDG No JC20060 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20160511

	Field Sample ID	Lab Sample ID		Method Blank (mg/l)	Sample Result	Validation Sample Result (mg/l)		 NJDEP Validation Footnote
	156-MW3-20160511	JC20060-1	CHROMIUM (HEXAVALENT)	U	0.020	0.020	0.010	
Ī	156-MW9-20160511	JC20060-3	CHROMIUM (HEXAVALENT)	U	0.031	0.031	0.010	

Note: A "U" under Method Blank column indicates a nondetect result.

AECOM Page 2 of 2

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 - Metro Towers

Sampling Date May 11, 2016

Lab Name/ID SGS/Accutest, Dayton, NJ

SDG No JC20060 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20160511

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Laboratory Sample Result (ug/l)	Validation Sample Result (ug/l)	RL (ug/l)	Quality Assurance Decision	NJDEP Validation Footnote
156-MW10-20160511	JC20060-5A	ANTIMONY	U	5.1B	5.1	6.0	Qualify	1
156-MW8-20160511	JC20060-2A	CHROMIUM	U	472	472	10		
156-MW9-20160511	JC20060-3A	CHROMIUM	U	1.3B	1.3	10	Qualify	1
156-MW10-20160511	JC20060-5A	CHROMIUM	U	42.2	42.2	10		
156-MW3-20160511	JC20060-1A	NICKEL	U	0.80B	0.80	10	Qualify	1
156-MW8-20160511	JC20060-2A	NICKEL	U	4.1B	4.1	10	Qualify	1
156-MW10-20160511	JC20060-5A	NICKEL	U	17.4	17.4	10		
156-MW3-20160511	JC20060-1A	VANADIUM	U	2.6B	2.6	50	Qualify	1
156-MW8-20160511	JC20060-2A	VANADIUM	U	6.8B	6.8	50	Qualify	1
156-MW9-20160511	JC20060-3A	VANADIUM	U	1.5B	1.5	50	Qualify	1
156-MW10-20160511	JC20060-5A	VANADIUM	U	10.9B	10.9	50	Qualify	1

Note: A "U" under Method Blank column indicates a nondetect result.

NJDEP Laboratory Footnote

1. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.

Attachment B

Data Validation Report Form

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation: Corr - Correlation Coefficient.

Χ

samples?

3) Was a sample spiked at the frequency of 1 per batch or 20

ITEM	YES	NO	N/A	COMMENTS
Post Digestion Spike			х	
1) Post Digestion Spike %R criteria met? (85-115%R).			х	
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	
Sample Duplicate Data Included in Lab Package?	х			JC20060-5
1) RPD criteria met? (RPD ≤20%) if both results are >4x RL or absolute difference ≤RL if either or both results are <4xRL.	Х			
2) Was a sample duplicate analyzed at the frequency of 1 per batch or 20 samples?	х			
Was a Laboratory Control Sample (LCS) Included in Lab Package?	х			
1) %R criteria met? (80-120%R).	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	х			
Were any Field Duplicate samples submitted with this SDG?		х		
1) RPD criteria met? (RPD ≤20%) if both results are >4x RL or absolute difference ≤RL if either or both results are <4xRL.			x	
Were all sample quantitation and reporting requirements met?	х			
Were all solid samples reported with percent solids > 50%?			х	NA -aqueous samples
2) Were any samples analyzed or reported with dilutions?		х		
Miscellaneous Items				
1) For soils by 7196A, was the pH within a range of 7.0-8.0?			х	
2) For soils by 7199, was the pH within a range of 9.0-9.5?			х	
3) For aqueous by 7196A, was the pH with a range of 1.5-2.5?	х			
4) For soils (3060A), was the digestion temperature 90-95C for at least 60 minutes?			х	
5) For 7199, was each sample injected twice and was the RPD <20?			х	

AECOM Page 4 of 5

SDG#: JC20060				
	x - concentration	y - response		
Batch: GN45715				
Cr+6 ICAL 05/11/2016	0	0.001		
Aqueous	0.01	0.01		
(p. 209 of data pkg)	0.05	0.047		
	0.1	0.097		
	0.3	0.281		
	0.5	0.469		
	0.8	0.756		
	1	0.935		(p. 209 of data
				(p. 209 oi data pkg)
AECOM Calculated Intercept	0.0011	OK	Reported intercept	0.0011
AECOM Slope	0.9373	OK	Reported Slope	0.9373
AECOM Calculated r	0.99998	OK	Reported r	0.99998
LCS calculation	GN45715-B1 pgs	. 205,209		
Background Absorbance	0			
Total absorbance	0.146			
Total absorbance - background	0.146			
Instrument Concentration	0.155			
Sample volume (mL)	50			
Final Volume (mL)	50			
Dilution Factor	1			1
AECOM Calculated LCS Result (mg/L)	0.15	OK	Reported Result (mg/L)	0.15
%R = Found/True*100	pgs. 205,209			
True Value (mg/L)	0.15			
AECOM Calculated %R	103.0	OK, rounding	Reported %R	100.0
		, ,	1	
MS calculation	GN45715-S3 (JC2	20060-5) pgs. 20°	7, 209	
Background reading	0.042			
Total absorbance	0.169			
Total absorbance - background	0.127			
Instrument Concentration	0.1343			
Sample volume (mL)	50			
Final Volume (mL)	50			
Percent solids	1			
Dilution Factor	1	014	D (1D 1) (11)	0.40
AECOM Calculated MS Result (mg/L)	0.13	OK	Reported Result (mg/L)	0.13
%R = Found/True*100	GN45715-S3 (JC2	20060-5) pgs. 20°	7, 209	
True Value (mg/L)	0.15	, 10		
Native concentration (mg/L)	0			
radive concentration (IIIU/L)	90	OK rounding	Reported %R	87
AECOM%R				
AECOM%R				
AECOM%R Reporting Limit	[156-MW10-2016	0511] (JC19912-1	1) pgs. 17, 130	
AECOM%R Reporting Limit Low Standard	0.01	0511] (JC19912-1	1) pgs. 17, 130	
AECOM%R Reporting Limit Low Standard Initial volume (mL)	0.01 50	0511] (JC19912-1	1) pgs. 17, 130	
AECOM%R Reporting Limit Low Standard Initial volume (mL) Final volume (mL)	0.01 50 50	0511] (JC19912-1	1) pgs. 17, 130	
AECOM%R Reporting Limit Low Standard Initial volume (mL) Final volume (mL) Percent solids	0.01 50 50 1	0511] (JC19912-1	1) pgs. 17, 130	
AECOM%R Reporting Limit Low Standard Initial volume (mL) Final volume (mL)	0.01 50 50	0511] (JC19912-1	1) pgs. 17, 130 Reported RL (mg/L)=	0.010

AECOM Page 5 of 5

Sample Calculations

	[156-MW10-20	0160511] (JC1	9912-1) pgs. 17, 130	
Background reading	0.0	44		
Total absorbance	0.0	45		
Total absorbance - background	0.0	01		
Instrument Response	0.0	00		
Sample volume (mL)	;	50		
Final Volume (mL)	;	50		
Percent solids		1		
Dilution Factor		1		
AECOM Calculated Result (mg/L)	0.0039U	OK	Reported Result (mg/L)	0.0039U

Client Name: PPG Industries			Project Number: 60493065.20				
Site Location: PPG Site 156 - Metro Towers			Project Manager: Nanette Conrey				
Laboratory: SGS/Accutest, Dayton, NJ		Т	уре с	of Validation: Limited			
Laboratory Job No: JC20060		С	Date C	checked: 06/30/2016			
Validator: Dawn Brule		F	eer: l	Mary Kozik			
ITEM	YES	NO	N/A	COMMENTS			
Sample results included?	Х						
Reporting Limits met project requirements?	Х						
Field I.D. included?	Х						
Laboratory I.D. included?	Х						
Did data package sample IDs match sample IDs on COC?	Х						
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	Х						
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	х						
Sample matrix included?	Х						
Sample receipt temperature 2-6°C?	Х			4.2° C			
Signed COCs included?	Х						
Date of sample collection included?	Х						
Date of sample digestion included?	Х						
Date of analysis included?	Х						
Holding time met QC criteria? (Metals -180 days from sample collection; Mercury - 28 days from sample collection. If HT exceeded by 10 days R all results.	x						
Method reference included?	Х						
Laboratory Case Narrative included?	Х						

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?	х			JC20060-1A (2x TI)
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) -Blank plus 1 standard? If no, reject (R) data.			Х	
3) Hg (7470/7471) -Blank plus 5 standards? If no, reject (R) data.			x	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			х	N/A for Limited Validation
Analyzed immediately after initial calibration? If no, reject (R) data.			х	
2) %R criteria met? (90-110%). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89%, and R all data for affected analyte(s) if %R <80% or >120%.			х	
3) Spot check ICV/ICCS results for several analytes.			Х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
Analyzed immediately after each ICV/ICC/CB and after every 10 samples? If no, reject (R) data.			x	
2) CCS and CCV from independent source and at mid-level of calibration curve. If no, reject (R) data.			х	
3) %R criteria met? (90-110%R). If no, J positive results for affected analyte(s) if %R between 80-89% and 111-120% and indicate bias; UJ non-detect results for affected analyte(s) if %R between 80-89% and R all data for affected analyte(s) if %R <80% or >120%.			х	
4) Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, Tl by 6010; 70-130% all others. If no, refer to ILM05.4 NJ SOP 5.A.2 for actions.			х	
Calibration Blanks			х	N/A for Limited Validation
Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples? If no, reject (R) data.			х	

ITEM	YES	NO	N/A	COMMENTS
2) Absolute value <3xIDL? If no, -if sample result <10x CB result, qualify affected analyte(s) in associated samples with CB; -if sample result >10xCB result, no qualification.			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? If no, reject (R) data, except no aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples? If - MB 1/25, J sample results from 21-25; -MB >1/25, R sample results after 25th sample.	х			
3) MB results nondetect? If no, -sample result <3xMB, negate UB; -sample result>3xMB but <10xMB, JB; -sample result >10xMB, no qualification.	х			<mdl< td=""></mdl<>
4) Negative MB result reported? If yes, -Positive sample result<10xMB, qualify estimated, biased low (J); -Non-detect sample result, qualify UJ, may be false non-detect.		х		<-mdl
Field Blanks/Equipment Blanks Included in Lab Package?	х			
The state of	х			
ICP Interference Check Sample (ICS) included in Lab Package?		x		N/A for Limited Validation
1) Analyzed at beginning of analytical run? If no, reject (R) data.			х	
2) %R criteria met? (80-120%) If no, %R>120%, no qualification if sample result non-detect; %R between 121-150%, J positive results, biased high; %R between 50-79%, J/UJ results, biased low; %R<50% or >150%, reject (R) result			х	
3) Spot check accuracy of %Rs			х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	х			Batch QC. QC not evaluated.
1) MS/MSD %R (75-125%R) and RPD (+20%) criteria met? - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; -%R<75% J/UJ for affected analyte(s) for all samples in the same batch/SDG; - RPD outside +20% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
3) Was the MS performed on a site sample?			х	
4) Was the MS performed on a FB/EB or TB? If yes, J all sample data.			х	
Post Digestion Spike			Х	
1) %R criteria met? (75-125%R) - %R>125% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs; - %R<75% J/UJ affected analyte(s) for all samples in the same batch/SDG.			х	
2) Was the spike performed on a FB/EB or TB? If yes, J all sample data.			Х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			Х	
Laboratory Duplicate Data Included in Lab Package?		х		MSD
Aqueous - If RPD is >20% but <100% and sample and duplicate results are >5x the QL, estimate (J) results >the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	Batch QC. QC not evaluated.
Soil - If RPD is >35% but <120% and sample and duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			х	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	Х			
1) LCS %R criteria met? (80-120%R). If no, J/UJ all affected analytes(s) for all samples in the same batch/SDG.	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples? If no, J/UJ affected analyte(s) for all samples in the same batch/SDG.	х			
Serial Dilution	Х			Batch QC. QC not evaluated.
1) %D (<10%R) criteria met? - If analyte concentration >25xIDL (7000) or >10xIDL (6010) and %D >10% J positive results for affected analyte(s) for all samples in the same batch/SDG, accept NDs.			х	
2) Was the frequency 1/batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
3) Was a site sample used?		Х		
4) Was a FB/EB or TB used? If yes, J all sample data.			х	
5) Spot check accuracy of %Ds.			х	
Field Duplicate Data included in Lab Package?		Х		
Aqueous - If RPD is >20% but <100% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >100%, reject R results >/= the QL If sample and/or duplicate is <5x the QL and absolute difference is > the QL, estimate (J) positive results <5x the QL and nondetects (UJ) If absolute difference is >2x the QL, reject R non-detects and positive results <5x the QL.			x	
Soil - If RPD is >35% but <120% and sample and field duplicate results are >5x the QL, estimate (J) results > the QL If RPD is >120%, reject results > the QL If sample and/or duplicate is <5x the QL and absolute difference is >2x the QL, estimate (J) positive results <5x QL and nondetects (UJ) If absolute difference is >4x the QL, reject nondetects and positive results <5x QL.			x	
Percent Solids data included in Lab Package?			х	Aqueous samples
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			х	



Data Validation Report

Project:	PPG Site 156 – Metro Towers 2017 Supplemental RI					
Laboratory:	SGS/Accutest, Dayton, NJ					
Laboratory Job No.:	JC55900					
Analysis/Method:	Hexavalent Chromium SW846 3060A/7196A CCPW Metals SW846 6010					
Validation Level:	Full (Hexavalent Chromium) Limited (CCPW Metals)					
Site Location/Address:	PPG Site 156 - Metro Towers, Jersey City, NJ					
AECOM Project No:	60493065.22					
Prepared by: Sharon M	cKechnie /AECOM	Completed on: 01/04/2018				
Reviewed by: Mary Kozi	k/AECOM File Name: JC55900_20)18-01-04_DVReport-F				
Introduction						

Introduction

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP validation Standard Operating Procedure(s) (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium for USEPA SW-846 Method 3060A, USEPA SW-846 Method 7196A;
- ICP-AES Data Validation, SOP No. HW-3a Revision 0 (July 2015).

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.
- UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.
- JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.

AECOM 2

R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.

RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on November 21, 2017 as part of the 2017 Supplemental RI sampling at PPG Site 156 - Metro Towers, Jersey City, NJ. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction*
156-FB-20171121 (Equipment Blank)	JC55900-4	Aqueous	Hexavalent Chromium
156-FB-20171121 (Equipment Blank)	JC55900-4A	Aqueous	CCPW Metals
156-MW8A-20171121	JC55900-1	Ground Water	Hexavalent Chromium
156-MW8A-20171121	JC55900-1A	Ground Water	CCPW Metals
156-MW8A-20171121X (Field Duplicate of 156-MW8A-20171121)	JC55900-2	Ground Water	Hexavalent Chromium
156-MW8A-20171121X (Field Duplicate of 156-MW8A-20171121)	JC55900-2A	Ground Water	CCPW Metals
156-MW8B-20171121	JC55900-3	Ground Water	Hexavalent Chromium
156-MW8B-20171121	JC55900-3A	Ground Water	CCPW Metals
*CCPW Metals: Antimony, Chromium, Nickel, That	llium, Vanadium		

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers, Jersey City, NJ and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The Eh/pH diagram included in the original issue of the laboratory data package was incomplete. A revised report was issued on 1/4/2018with a complete Eh/pH diagram.

Quality control (QC) issues identified during validation are discussed below. Refer to the Soil Target Analyte Summary Hit List(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

MS Results

Sample 156-MW8A-20171121 was selected for the matrix spike (MS) analysis and used for supporting data quality recommendations. The soluble MS recovery was 106.7%, which met the quality control (QC) criteria of 85-115%R. Therefore, all hexavalent chromium results were accepted without qualification.

All the samples were tested for pH and oxidation reduction potential (ORP) and plotted on an Eh/pH phase diagram. From this diagram, the source sample for the matrix spike analysis was plotted below the phase change line, indicating a reducing potential within the sample matrix incapable of supporting hexavalent chromium.

AECOM 3

CCPW Metals

Equipment Blanks

Nickel was detected in the equipment blank (EB) associated with the soil samples in this data set, at a concentration above the method detection limit (MDL), but below the reporting limit (RL). Nickel was detected in samples 156-MW8A-20171121 and 156-MW8A-20171121X at a concentration greater than three but less than ten times the EB concentration; therefore, the nickel results in both samples were qualified as estimated (J).

Sample Results

Reported results (flagged B by the laboratory) that were less than the RL, but greater than or equal to the MDL, are approximate values and have been qualified as estimated (J).

Data Quality and Usability

In general, all data appear to be valid and may be used for decision-making purposes. Qualified and detected results are presented in Attachments A and B.

All hexavalent chromium results were accepted without qualification.

The nickel results in samples 156-MW8A-20171121 and 156-MW8A-20171121X were qualified as estimated due to equipment blank contamination.

Sample results reported between the MDL and RL are usable as estimated values with an unknown directional bias.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hit List(s)

AECOM 1 of 2

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 – Metro Towers 2017 Supplemental RI

Sampling Date November 21, 2017

Lab Name/ID SGS/Accutest, Dayton, NJ

SDG No JC55900 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20171121

No positive results and no qualified nondetects

AECOM 2 of 2

Aqueous Target Analyte Summary Hit List (Metals)

Site Name PPG Site 156 – Metro Towers 2017 Supplemental RI

Sampling Date November 21, 2017
Lab Name/ID SGS/Accutest, Dayton, NJ

SDG No JC55900 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20171121

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Laboratory Sample Result (ug/l)	Validation Sample Result (ug/l)	RL (ug/l)	Assurance	NJDEP Validation Footnote
156-FB-20171121	JC55900-4A	NICKEL	U	1.3B	1.3J	10	Qualify	2
156-MW8A-20171121	JC55900-1A	CHROMIUM	U	0.90B	0.90J	10	Qualify	2
156-MW8A-20171121	JC55900-1A	NICKEL	U	10.6	10.6J	10	Qualify	1
156-MW8A-20171121	JC55900-1A	VANADIUM	U	1.7B	1.7J	50	Qualify	2
156-MW8A-20171121X	JC55900-2A	CHROMIUM	U	1.4B	1.4J	10	Qualify	2
156-MW8A-20171121X	JC55900-2A	NICKEL	U	10.9	10.9J	10	Qualify	1
156-MW8A-20171121X	JC55900-2A	VANADIUM	U	2.5B	2.5J	50	Qualify	2
156-MW8B-20171121	JC55900-3A	CHROMIUM	U	1.9B	1.9J	10	Qualify	2
156-MW8B-20171121	JC55900-3A	VANADIUM	U	22.8B	22.8J	50	Qualify	2

Note: A "U" under Method Blank column indicates a nondetect result.

NJDEP Laboratory Footnote

- 1. The value reported is greater than three (3) times but less than ten (10) times the value in the trip/field blanks and is considered "real". However, the reported value must be quantitatively qualified "J" due to trip/field blank contamination.
- 2. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated.

Attachment B

Data Validation Report Form

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation: Corr - Correlation Coefficient.

Laboratory Case Narrative included?

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	Х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	Х			
2) Correlation coefficient of >0.995 (7196A) or >0.999 (7199)	Х			
3) Calibrate daily or each time instrument is set up.	Х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	X			
1) %R criteria met? (90 - 110%)	Х			
2) Correct frequency of one per every 10 samples	Х			
3) CCS and QCS from independent source and at mid-level of calibration curve	X			
Calibration Blanks	Х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	Х			
2) Absolute value should not exceed MDL.	х			
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	х			
1) Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	Х			
1) Eh and pH data was included and plotted for all samples?		X*		* See memo
Soluble Matrix Spike Data Included in Lab Package?	Х			JC55900-1, soluble %R 106.7%. No table.
1) Soluble Matrix %R criteria met? (85-115%R).	Х			
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?			х	As an aqueous sample, JC55900-1 was spiked at 0.150 mg/l. The data were not affected.
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	Х			
Insoluble Matrix Spike Data Included in Lab Package?			х	NA for aqueous samples
1) Insoluble Matrix %R criteria met? (85-115%R).			х	
2) Was the spike concentration around 400 to 800 mg/Kg?			х	
Post Spike		х		None reported
1) Post Spike %R criteria met? (85-115%R).			х	
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?			х	

ITEM	YES	NO	N/A	COMMENTS
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	
Sample Duplicate Data Included in Lab Package?	х			JC55900-1
1) RPD criteria met? (RPD ≤ 20%) if both results are ≥4x RL or absolute difference ±RL if either or both results are <4xRL	х			Sample and duplicate all nondetect. No table.
Was a sample duplicate analyzed at the frequency of 1 per batch or 20 samples?	х			
Was a Laboratory Control Sample (LCS) Included in Lab Package?	х			
1) %R criteria met? (80-120%R).	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	х			
Were any Field Duplicate samples submitted with this SDG?	х			156-MW8A-20171121/156-MW8A-20171121X
1) Were Field duplicate RPD criteria met? (RPD≤20% if both sample results ≥4xR, or absolute difference ±RL if either or both results are <4xRL.	х			Both ND, OK. No table
Were all sample quantitation and reporting requirements met?	х			
Were all solid samples reported with percent solids > 50%?			х	NA aqueous samples
2) Were any samples analyzed or reported with dilutions?		х		
Miscellaneous Items	х			
1) For soils by 7196A, was the pH within a range of 7.0-8.0?			х	
2) For soils by 7199, was the pH within a range of 9.0-9.5?			х	
3) For aqueous by 7196A, was the pH with a range of 1.5-2.5?	х			
4) For soils (3060A), was the digestion temperature 90-95C for at least 60 minutes?			х	
5) For 7199, was each sample injected twice and was the RPD ≤20?			х	
Total Chromium concentration >hexavalent chromium in associated samples?	х			All OK, no table.

		T	\neg	
SDG#: JC55900, Method 7196	x - concentration	y - response		
Batch: GN72970				
Cr+6 ICAL - 11/21/2017	0	0		
Groundwater	0.01	0.008		
(p 319 of data pkg)	0.05	0.047		
	0.1	0.097		
	0.3	0.257		
	0.5	0.461		
	0.8	0.728		
	1	0.925		
				(p 319 of data pkg)
AECOM Calculated Intercept	-0.0019	OK	Reported intercept	-0.0019
AECOM Slope	0.9195	OK	Reported Slope	0.9195
AECOM Calculated r	0.99975	OK	Reported r	0.99975
LOO aslantation	ON 700 70 D4	D 045 040		
LCS calculation	GN72970-B1	P.315,319		
Background absorbance	0			
Sample absorbance	0.142			
LCS Soluble Instrument Response	0.142			
Instrument Concentration (mg/L)	0.156			
Sample aliquot	0.05	NA: standards same	and samples treated the	
Final volume	0.05	Same		
Dilution Factor	1			
			Reported Result	
AECOM Calculated LCS Result (mg/L)	0.16	OK	(mg/L)	0.16
%R = Found/True*100	GN72970-B1	P.315,319		
				Reported in raw
True Value (mg/L)	0.15			data as 104% p.320
AECOM Calculated %R	104	OK	Reported %R	106.7
ALGON Galculated 7010	104	Oit	reported 7014	100.7
MS calculation	GN72970-S1	P.317,319	JC55900-1	
Background reading	0.003			
Total absorbance	0.146			
Total absorbance - background	0.143			
Instrument Concentration (mg/L)	0.1575			
			and samples treated the	
Sample aliquot	0.05	same		
Final volume	0.05			
Dilution Factor	1		Reported Result	
AECOM Calculated MS Result (mg/L)	0.16	OK	(mg/L)	0.16
%R = Found/True*100	GN72970-S1	P.317,319	JC55900-1	
True Value (mg/L)	0.150	,		
Native concentration (mg/L)	0.000			
AFOON O-1	167.0		D	460 =

105.0

Reported %R

106.7

Reporting Limit	JC55900-1	P.15	156-MW8A- 20171121	
Low Standard	0.01			
Initial volume (L) Final volume (L)	0.05 0.05	NA: standards ar same	nd samples treated the	
Dilution Factor AECOM Calculated Reporting Limit	0.010	OK	Reported RL (mg/L)=	0.010
ALCOM Calculated Reporting Limit	0.010	OK	Reported RL (mg/L)=	0.010
Sample Calculations	JC55900-1	P.8,319	156-MW8A- 20171121	
Background reading	0.003			
Total absorbance	0.004			
Total absorbance - background	0.001			
Instrument Response (mg/L)	0.003			
Sample volume (L) Final Volume (L)	0.05 0.05	NA: standards ar same	nd samples treated the	
Dilution Factor	0.03			
AECOM Calculated Result (mg/L)	0.003	OK, Reported as Nondetect	Reported Result (mg/L)	0.0081 U

Client Name: PPG Industries			Project Number: 60493065				
Site Location: PPG Site 156 – Metro Towers 2017 Supplemental RI Jersey City, NJ			Project Manager: Bill Spronz				
Laboratory: SGS/Accutest, Dayton, NJ			e of \	Validation: Limited			
Laboratory Job No: JC55900		Dat	e Che	ecked: 01/04/2018			
Validator: Sharon McKechnie		Pee	er: Ma	ıry Kozik			
ITEM	YES	NO	N/A	COMMENTS			
Sample results included?	Х						
Reporting Limits met project requirements?	Х						
Field I.D. included?	Х						
Laboratory I.D. included?	Х						
Did data package sample IDs match sample IDs on COC?	Х						
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	Х						
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	Х						
Sample matrix included?	Х						
Sample receipt temperature 2-6°C?	Х						
Signed COCs included?	Х						
Date of sample collection included?	Х						
Date of sample digestion included?			Х	Aqueous samples			
Date of analysis included?	Х						
Holding time met QC criteria? (Metals -180 days from sample collection; Mercury - 28 days from sample collection. If HT exceeded by 10 days R all results.	х						
Method reference included?	Х						
Laboratory Case Narrative included?	Х						

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?	х			Up to 5X
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
1) Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) -Blank plus 1 standard?			х	
3) Hg (7470/7471) -Blank plus 5 standards			х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			x	N/A for Limited Validation
1) Analyzed immediately after initial calibration? If no, reject (R) data.			х	
2) %R criteria met? (90-110%).			х	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			x	N/A for Limited Validation
1) Analyzed immediately after each ICV/ICC/CB and after every 10 samples?			х	
2) CCS and CCV from independent source and at mid-level of calibration curve.			x	
3) %R criteria met? (90-110%R).			Х	
4) Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	N/A for Limited Validation
1) %R criteria met? -			х	
Calibration Blanks			х	N/A for Limited Validation
1) Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples?			x	
2) Absolute value <3xIDL?			х	
Method Blank Included in Lab Package?	х			
1) Method blank analyzed with each preparation batch or every SDG, or 1/20 samples?	х			
2) Method blank analyzed 1/20 samples?	х			
3) MB results nondetect?	х			
4) Negative MB result reported?		х		

ITEM	YES	NO	N/A	COMMENTS
Field Blanks/Equipment Blanks Included in Lab Package?	Х			
1) FB/EB result non-detect?		х		See table
ICP Interference Check Sample (ICS) included in Lab Package?			x	N/A for Limited Validation
1) Analyzed at beginning of analytical run?			Х	
2) %R criteria met? (80-120			Х	
3) Spot check accuracy of %Rs			Х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	x			JC55900-1A
1) MS/MSD %R (75-125%R) and RPD (20%) criteria met?	х			
2) Was a sample spiked at the frequency of 1/batch or 20 samples?	x			
3) Was the MS performed on a site sample?	х			
4) Was the MS performed on a FB/EB or TB?		х		
Post Digestion Spike			х	
1) %R criteria met? (75-125%R)			Х	
2) Was the spike performed on a FB/EB or TB?			Х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			х	
Laboratory Duplicate Data Included in Lab Package?		х		
Aqueous – Is RPD ≤ 20%?			Х	
Soil – Is RPD ≤ 35%?			Х	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	x			
1) LCS %R criteria met? (80-120%R).	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	х			
Serial Dilution			х	N/A for Limited Validation
1) %D (<10%R) criteria met? -			х	
2) Was the frequency 1/batch or 20 samples?			х	
3) Was a site sample used?			х	
4) Was a FB/EB or TB used?			Х	

ITEM	YES	NO	N/A	COMMENTS
5) Spot check accuracy of %Ds.			Х	
Field Duplicate Data included in Lab Package?	Х			156-MW8A-20171121/156-MW8A-20171121X
Aqueous - Is RPD is <20%?		Х		No action; see table
Soil - Is RPD is <35%?			Х	
Percent Solids data included in Lab Package?			Х	
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)			Х	
Total Chromium concentration >hexavalent chromium in associated samples?	х			

AECOM Page 5 of 5

Blanks

Analyte	Result	3X	10X	Sample Actions
Equipment Blank	(ug/L)	(ug/L)	(ug/L)	
				JC55900-1A, JC55900-2A >3X but <10X EB; Estimate J
Nickel	1.3	3.9	13	JC55900-3A Nondetect; Accept

Field Duplicate

Analyte	Sample	Result	RL	Duplicate	Dup Result	Dup RL	Units	Actions
CHROMIUM	156-MW8A-20171121	0.9B	10	156-MW8A-20171121X	1.4B	10	ug/l	Both results less than 5X
VANADIUM	156-MW8A-20171121	1.7B	50	156-MW8A-20171121X	2.5B	50	ug/l	RL, absolute difference less than the RL; Accept



Data Validation Report

Project:	wers 2017 Supplemental RI					
Laboratory:	SGS/Accutest, Dayton, N	IJ				
Laboratory Job No.: JC57819						
Analysis/Method:		Hexavalent Chromium SW846 3060A/7196A Antimony, Chromium, Nickel, Thallium, and Vanadium SW-846 8050B/6010C				
Validation Level:	Full (Hexavalent Chromium) Limited (Metals)					
Site Location/Address:	PPG Site 156 - Metro Towers, Jersey City, New Jersey					
AECOM Project No:	60493065					
Prepared by: Charlene Livingston Flint /AECOM		Completed on: 01/05/2018				
Reviewed by: Mary Kozik/AECOM Introduction		File Name: JC57819_2018-01-05_DVReport-F				

The data were reviewed in accordance with the FSP-QAPP and the following NJDEP validation Standard Operating Procedures (SOP):

- NJDEP Office of Data Quality SOP 5.A.10, Rev 3 (September 2009), SOP for Analytical Data Validation of Hexavalent Chromium - for USEPA SW-846 Method 3060A, USEPA SW-846 Method 7196A and USEPA SW-846 Method 7199;
- NJDEP Office of Data Quality SOP 5.A.16, Rev 1 (May 2002), Quality Assurance Data Validation of Analytical Deliverables for Inorganics (based on USEPA SW-846 Methods)

The results of quality control data analyzed with site samples were used to assess the overall reliability of the data. The following qualifiers were used to identify data quality issues:

- U: Indicates the analyte was not detected in the sample above the sample reporting limit.
- J: Indicates the result was an estimated value; the associated numerical value was an approximate concentration of the analyte in the sample.
- J+: Indicates the analyte was positively identified; the associated numerical value is an estimated quantity with a potential high bias.
- J-: Indicates the analyte was positively identified; the associated numerical value is an estimated quantity with a potential low bias.
- UJ: Indicates the analyte was not detected above the reporting limit and the reporting limit was approximate.

AECOM 2

UB: The analyte concentration is less than or equal to three (3) times the concentration in the associated method/prep blank. The presence of the analyte in the sample is negated (UB) due to laboratory contamination.

- JB: The analyte concentration is greater than three (3) times, but less than or equal to ten (10) times the concentration in the associated method/prep blank. The presence of that analyte in the sample is considered "real". The concentration is quantitatively qualified (JB) due to method blank contamination.
- R: The sample result was rejected due to serious deficiencies; the presence or absence of the analyte could not be confirmed.
- RA: The sample result was rejected but is still considered usable.

Sample Information

The samples listed below were collected by AECOM on December 21, 2017 as part of the PPG Site 156 - Metro Towers 2017 Supplemental RI at Jersey City, New Jersey. Only the samples and parameters listed below were validated.

Field ID	Laboratory ID	Matrix	Fraction
156-FB-20171221 (Equipment Blank)	JC57819-4, -4A	Aqueous	Hexavalent Chromium and Metals*
156-MW8A-20171221	JC57819-1, -1A	Ground Water	Hexavalent Chromium and Metals*
156-MW8A-20171221X (Field Duplicate of156-MW8A-20171221)	JC57819-2, -2A	Ground Water	Hexavalent Chromium and Metals*
156-MW8B-20171221	JC57819-3, -3A	Ground Water	Hexavalent Chromium and Metals*
*Metals (Sb, Cr, Ni, Tl, V)		•	•

The samples were collected following the procedures detailed in the Work Order for PPG Site 156 - Metro Towers 2017 Supplemental RI at Jersey City, New Jersey and the Field Sampling Plan/Quality Assurance Project Plan for Non-Residential and Residential Chromium Sites Hudson County, New Jersey (December 2011).

General Comments

The data package was complete. Quality control (QC) issues identified during validation are discussed below. Refer to the Target Analyte Summary Hit List(s) in Attachment A for a listing of all detected results, qualified results, and associated qualifications, where applicable. The nonconformances for each section discussed below are presented in Attachment B.

Hexavalent Chromium

There were no QC nonconformances noted during validation, thus the data were accepted as reported by the laboratory.

AECOM 3

MS Results

Sample 156-MW8A-20171221 (JC57819-1) was selected for the matrix spike (MS) analysis associated with the samples in this SDG and was used for supporting data quality recommendations. The aqueous MS recovery was 100%; which met the quality control criteria of 85-115%. No data qualification was required on the basis of spike recoveries.

TAL Metals

Laboratory Blanks/Equipment Blanks

Nickel was reported in the Equipment Blank (EB) 156-FB-20171221, at a concentration greater than the method detection limit (MDL) but less than the reporting limit (RL). The associated samples 156-MW8A-20171221 (JC57819-1A) and 156-MW8A-20171221X (JC57819-2A) reported nickel at concentrations less than three times the EB; therefore, nickel was negated (UJ) in these samples. Nickel was not detected in sample 156-MW8B-20171221 (JC57819-3A); therefore, no qualification was necessary for this sample.

Sample Results

Reported results (flagged B by the laboratory) that were less than the reporting limit (RL), but greater than or equal to the method detection limit (MDL) are approximate values that have been qualified as estimated (J).

The reporting limits for two nondetect thallium results were above the NJDEP Specific Ground Water Quality Criteria due to elevated levels of an interfering element. These nondetect results are usable for project objectives as elevated reporting limits. Refer to the Target Analyte Summary Hitlist(s) in Attachment A and the nonconformance table(s) in Attachment B for a list of impacted sample results.

Data Quality and Usability

In general, these data appear to be valid and may be used for decision-making purposes. No data were rejected. Qualified results, if applicable, are presented in Attachments A and B.

The nondetect results for thallium in two samples as presented in Attachment B, exceeded the Ground Water Reporting Limits Criteria; therefore, the nondetect results may not meet project objectives.

Sample results reported between the MDL and RL are usable as estimated values with an unknown directional bias.

ATTACHMENTS

Attachment A: Target Analyte Summary Hit List(s)

Attachment B: Data Validation Report Form

Attachment A

Target Analyte Summary Hit List(s)

AECOM Page 1 of 1

Aqueous Target Analyte Summary Hit List (Hexavalent Chromium)

Site Name PPG Site 156 - Metro Towers 2017 Supplemental RI

Sampling Date December 21, 2017
Lab Name/ID SGS/Accutest, Dayton, NJ

SDG No JC57819 Sample Matrix Aqueous

Trip Blank ID NA

Field Blank ID 156-FB-20171221

ı	Field Sample ID	Lab Sample ID	Analyte	Method Blank (mg/l)	Laboratory Sample Result (mg/l)	Validation Sample Result (mg/l)	RL (mg/l)	Assurance	NJDEP Validation Footnote
1	No positive results and no c	qualified non-detec	ts						

AECOM Page 1 of 1

Aqueous Target Analyte Summary Hit List (TAL Metals)

Site Name PPG Site 156 - Metro Towers 2017 Supplemental RI

Sampling Date December 21, 2017
Lab Name/ID SGS/Accutest, Dayton, NJ

SDG No JC57819 Sample Matrix Aqueous Trip Blank ID NA

Field Blank ID 156-FB-20171221

Field Sample ID	Lab Sample ID	Analyte	Method Blank (ug/l)	Laboratory Sample Result (ug/l)	Validation Sample Result (ug/l)	RL (ug/l)	Quality Assurance Decision	NJDEP Validation Footnote
156-MW8A-20171221	JC57819-1A	CHROMIUM	U	2.2B	2.2J	10	Qualify	1
156-MW8A-20171221X	JC57819-2A	CHROMIUM	U	2.4B	2.4J	10	Qualify	1
156-MW8B-20171221	JC57819-3A	CHROMIUM	U	4.5B	4.5J	10	Qualify	1
156-MW8A-20171221	JC57819-1A	NICKEL	U	9.2B	UJ	10	Negate	1,2
156-MW8A-20171221X	JC57819-2A	NICKEL	U	9.4B	UJ	10	Negate	1,2
156-FB-20171221	JC57819-4A	NICKEL	U	4.9B	4.9J	10	Qualify	1
156-MW8B-20171221	JC57819-3A	THALLIUM	U	9.7B	9.7J	10	Qualify	1
156-MW8A-20171221	JC57819-1A	VANADIUM	U	2.6B	2.6J	50	Qualify	1
156-MW8A-20171221X	JC57819-2A	VANADIUM	U	2.3B	2.3J	50	Qualify	1
156-MW8B-20171221	JC57819-3A	VANADIUM	U	41.8B	41.8J	50	Qualify	1

Note: A "U" under Method Blank column indicates a nondetect result.

NJDEP Laboratory Footnote

- 1. The reported result was greater than the MDL but less than the RL and qualified (J) as estimated by the laboratory.
- 2. The value reported is less than or equal to 3x the value in the trip/field blank. It is the policy of NJDEP-DPFSR to negate the reported value as due to probable foreign contamination unrelated to the actual sample. The end-user, however, is alerted that a reportable quantity of the analyte was detected.

Attachment B

Data Validation Report Form

Client Name: PPG Industries			Project Number: 60493065					
Site Location: PPG Site 156 - Metro Towers Supplemental RI, Jersey City, NJ	Site Location: PPG Site 156 - Metro Towers 2017 Supplemental RI, Jersey City, NJ			Project Manager: William Spronz				
Laboratory: SGS/Accutest, Dayton, NJ			Type of Validation: Full					
Laboratory Job No: JC57819			Date Checked: 01/05/2017					
Validator: Charlene Livingston Flint		F	Peer: I	Mary Kozik				
ITEM	YES	NO	N/A	COMMENTS				
Sample results included?	Х							
Reporting Limits met project requirements?	Х							
Field I.D. included?	Х							
Laboratory I.D. included?	Х							
Did data package sample IDs match sample IDs on COC?	х							
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	х							
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	х							
Sample matrix included?	Х							
Sample receipt temperature 2-6 °C?	Х			3 °C				
Signed COCs included?	Х							
Date of sample collection included?	Х							
Date of sample digestion included?			Х	Aqueous samples				
Holding time to digestion met criteria? (Soils -30 days from collection to digestion.)			х	Aqueous samples				
Date of analysis included?	Х							
Holding time to analysis met criteria? (Soils -168 hours from digestion to analysis; Aqueous - 24 hours from collection to analysis.	х							
Method reference included?	Х							
Laboratory Case Narrative included?	Х							
Definitions: MDL - Method Detection Limit; %R - P	ercent	Recov	very; R	L - Reporting Limit; RPD - Relative Percent				

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Initial calibration documentation included in lab package?	х			
1) Blank plus 4 standards (7196A) or blank plus 3 standards (7199)	х			
2) Correlation coefficient of =0.995 (7196A) or =0.999 (7199)	х			
3) Calibrate daily or each time instrument is set up.	Х			
Calibration Check Standard (CCS) for 7196A and Quality Control Sample (QCS) for 7199 Included in Lab Package?	х			
1) %R criteria met? (90 - 110%)	Х			
2) Correct frequency of one per every 10 samples	Х			
3) CCS and QCS from independent source and at mid- level of calibration curve	х			
Calibration Blanks	Х			
Analyzed prior to initial calibration standards and after each CCS/QCS?	х			
2) Absolute value should not exceed MDL.		х		CCB>MDL, SR ND No qualifications made.
Method Blank, Field Blanks and/or Equipment Blanks Included in Lab Package?	х			
1) Method blank analyzed with each preparation batch?	Х			
2) Absolute value should not exceed MDL.	Х			
Eh and pH Data	Х			
The sample of the sample	x			
Soluble Matrix Spike Data Included in Lab Package?	Х			JC57819-1
1) Soluble Matrix %R criteria met? (75-125%R).	Х			
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?			х	Aq sample. Spiked at 0.150 mg/l
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?	х			
Insoluble Matrix Spike Data Included in Lab Package?			х	
1) Insoluble Matrix %R criteria met? (75-125%R).			Х	
2) Was the spike concentration around 400 to 800 mg/Kg?			Х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	

ITEM	YES	NO	N/A	COMMENTS
Post Digestion Spike			х	
1) Post Digestion Spike %R criteria met? (85-115%R).			х	
2) Was the spike concentration 40 mg/Kg or twice the sample concentration?			х	
3) Was a sample spiked at the frequency of 1 per batch or 20 samples?			х	
Sample Duplicate Data Included in Lab Package?	Х			JC57819-1
1) RPD criteria met? (RPD ≤ 20%) if both results are ≥4x RL or absolute difference ±RL if either or both results are <4xRL	х			Both ND
2) Was a sample duplicate analyzed at the frequency of 1 per batch or 20 samples?	x			
Was a Laboratory Control Sample (LCS) Included in Lab Package?	x			
1) %R criteria met? (80-120%R).	Х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	x			
Were any Field Duplicate samples submitted with this SDG?	x			156-MW8A-20171221 (JC57819-1) & 156-MW8A-20171221X (JC57819-2)
1) Were Field duplicate RPD criteria met? (RPD≤20% if both sample results ≥4xR, or absolute difference ±RL if either or both results are <4xRL.	х			Both ND
Were all sample quantitation and reporting requirements met?	х			
1) Were all solid samples reported with percent solids > 50%?			х	Aq samples
2) Were any samples analyzed or reported with dilutions?		х		
Miscellaneous Items	Х			
1) For soils by 7196A, was the pH within a range of 7.0-8.0?			х	
2) For soils by 7199, was the pH within a range of 9.0-9.5?			х	
3) For aqueous by 7196A, was the pH with a range of 1.5-2.5?	х			
4) For soils (3060A), was the digestion temperature 90-95C for at least 60 minutes?			х	
5) For 7199, was each sample injected twice and was the			х	

AECOM DATA VALIDATION REPORT FORM - HEXAVALENT CHROMIUM ANALYSIS 7196 Page 4 of 6

RPD <u><</u> 20?			
Total Cr vs Cr6 Fraction Agreement Reviewed	Х		Total CR greater than total hexavalent chromium

SDG#: JC57819/ Method 7196	x - concentration	y - response		
Batch: GN74213				
Cr+6 ICAL 12/21/17	0	0		
Soil	0.01	0.009		
(p. 225 of data pkg)	0.05	0.045		
	0.1	0.08		
	0.3	0.271		
	0.5	0.481		
	0.8	0.749		
	1	0.939		
			D	(p. 225 of data pkg)
AECOM Calculated Offset	-0.0040	OK	Reported Offset	-0.0040
AECOM Slope	0.9442	OK	Reported Slope	0.9442
AECOM Calculated r	0.99979	OK	Reported r	0.99979
LCS calculation	GP74213-B1	P. 221,225		
Background Absorbance	0	1.221,223		
Total absorbance	0.137			
Total absorbance - background	0.137			
Instrument Concentration	0.137			
Sample weight (mg/l)	0.05			
Final Volume (L)	0.05			
Dilution Factor	0.03			
AECOM Calculated LCS Result (mg/l)	0.149	OK	Reported Result (mg/Kg)	0.15
ALCON Calculated LCG Result (High)	0.149	OK	Reported Result (mg/Rg)	0.15
%R = Found/True*100	GP74213-B1	P. 221,225		
%R = Found/True*100 True Value (mg/l)	GP74213-B1 0.15	P. 221,225		
		P. 221,225 OK	Reported %R	100
True Value (mg/l) AECOM Calculated %R	0.15	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation	0.15 100 GP74213-S1		Reported %R JC57819-1	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading	0.15 100 GP74213-S1 0.003	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance	0.15 100 GP74213-S1 0.003 0.141	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background	0.15 100 GP74213-S1 0.003 0.141 0.138	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05	OK	·	100
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05	OK P. 223,225	JC57819-1	
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05	OK	·	0.15
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1	OK P. 223,225 OK	JC57819-1 Reported Result (mg/l)	
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15	OK P. 223,225	JC57819-1	
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15	OK P. 223,225 OK	JC57819-1 Reported Result (mg/l)	
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l) Native concentration (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15 GP74213-S1 0.15	OK P. 223,225 OK P. 223,225	JC57819-1 Reported Result (mg/l) JC57819-1	0.15
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15	OK P. 223,225 OK	JC57819-1 Reported Result (mg/l)	
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l) Native concentration (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15 GP74213-S1 0.15	OK P. 223,225 OK P. 223,225	JC57819-1 Reported Result (mg/l) JC57819-1	0.15
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l) Native concentration (mg/l) AECOM%R	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15 GP74213-S1 0.15 0 100	OK P. 223,225 OK P. 223,225 OK	JC57819-1 Reported Result (mg/l) JC57819-1 Reported %R	0.15
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l) Native concentration (mg/l) AECOM%R Reporting Limit	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15 GP74213-S1 0.15 0 100	OK P. 223,225 OK P. 223,225 OK	JC57819-1 Reported Result (mg/l) JC57819-1 Reported %R	0.15
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l) Native concentration (mg/l) AECOM%R Reporting Limit Low Standard	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15 0.15 GP74213-S1 0.15 0 100 JC57819-1 0.01	OK P. 223,225 OK P. 223,225 OK	JC57819-1 Reported Result (mg/l) JC57819-1 Reported %R	0.15
True Value (mg/l) AECOM Calculated %R MS calculation Background reading Total absorbance Total absorbance - background Instrument Concentration Sample weight (mg/l) Final Volume (L) Dilution Factor AECOM Calculated MS Result (mg/l) %R = Found/True*100 True Value (mg/l) Native concentration (mg/l) AECOM%R Reporting Limit Low Standard Initial weight (mg/l)	0.15 100 GP74213-S1 0.003 0.141 0.138 0.1504 0.05 0.05 1 0.15 0.15 GP74213-S1 0.15 0 100 JC57819-1 0.01 0.05	OK P. 223,225 OK P. 223,225 OK	JC57819-1 Reported Result (mg/l) JC57819-1 Reported %R	0.15

Sample Calculations	JC57819-1	P. 8,225	156-MW8A-20171221	
Background reading	0.003			
Total absorbance	0.004			
Total absorbance - background	0.001			
Instrument Response	0.005			
Sample weight (mg/l)	0.05			
Final Volume (L)	0.05			
Dilution Factor	1			
AECOM Calculated Result (mg/l)	0.005	<mdl, nd<="" td=""><td>Reported Result (mg/l)</td><td>0.0081 U</td></mdl,>	Reported Result (mg/l)	0.0081 U

Client Name: PPG Industries			Project Number: 60493065				
Site Location: PPG Site 156 - Metro Towers Supplemental RI	s 2017	F	Project Manager: William Spronz				
Laboratory: SGS/Accutest, Dayton, NJ			Type of Validation: Full				
Laboratory Job No: JC57819		С	Date Checked: 01/05/2017				
Validator: Charlene Livingston Flint		F	eer: l	Mary Kozik			
ITEM	YES	NO	N/A	COMMENTS			
Sample results included?	Х						
Reporting Limits met project requirements?		Х		See table below			
Field I.D. included?	Х						
Laboratory I.D. included?	Х						
Did data package sample IDs match sample IDs on COC?	Х						
Did electronic data deliverable (EDD) sample IDs match COC sample IDs?	Х						
Did data package sample IDs match electronic data deliverable (EDD) sample IDs?	Х						
Sample matrix included?	Х						
Sample receipt temperature 2-6 °C?	Х			3 °C			
Signed COCs included?	Х						
Date of sample collection included?	Х						
Date of sample digestion included?			Х	Aqueous samples			
Date of analysis included?	Х						
Holding time met QC criteria? (Metals -180 days from sample collection; Mercury - 28 days from sample collection. If HT exceeded by 10 days R all results.	x						
Method reference included?	Х						
Laboratory Case Narrative included?	Х						
Definitions: MDL - Method Detection Limit: %R - P	ercent	Recov	erv. R	I - Reporting Limit: RPD - Relative Percent			

Definitions: MDL - Method Detection Limit; %R - Percent Recovery; RL - Reporting Limit; RPD - Relative Percent Difference; RSD - Relative Standard Deviation :Corr - Correlation Coefficient.

ITEM	YES	NO	N/A	COMMENTS
Sample dilutions?	х			Sample 156-MW8B-20171221 (JC57819-3) was diluted 5X.
Initial calibration documentation included in lab package?			х	N/A for Limited Validation
1) Calibrate daily or each time instrument is set up.			х	
2) ICP (6010) - plus 1 standard?			х	
3) Cn and Hg -Blank plus 5 standards?			х	
Initial Calibration Verification Standard (ICV) for ICP (6010) and Initial Calibration Check Standard (ICCS) for Hg (7470/7471) included in lab package?			х	N/A for Limited Validation
1) Analyzed immediately after initial calibration?			х	
2) %R criteria met? (90-110%).			х	
3) Spot check ICV/ICCS results for several analytes.			х	
Continuing Calibration Verification Standard (CCV) for ICP (6010) and Calibration Check Standard (CCS) for Hg (7470/7471) included in Lab Package?			х	N/A for Limited Validation
1) Analyzed immediately after each ICV/ICC/CB and after every 10 samples?			х	
2) CCS and CCV from independent source and at mid-level of calibration curve?			x	
3) %R criteria met? (90-110%R).			х	
4) Spot check CCV/CCS results for several analytes.			х	
Low Calibration Standard (CRI) included in Lab Package?			х	
1) %R criteria met? - 50-150% for Co, Mn, Zn, by ICP-MS; Pb, Tl by 6010; 70-130% all others			х	
Calibration Blanks			х	N/A for Limited Validation
Analyzed after daily calibration and after each ICV/ICC/CCV/CCS and after every 10 samples?			х	
2) Absolute value <3xIDL?			х	
Method Blank Included in Lab Package?	х			
Method blank analyzed with each preparation batch or every SDG, or 1/20 samples? No aqueous MB required for FB/EB if only soil samples were analyzed.	х			
2) Method blank analyzed 1/20 samples?	х			
3) MB results nondetect?	х			
4) Negative MB result reported?		х		

ITEM	YES	NO	N/A	COMMENTS
Field Blanks/Equipment Blanks Included in Lab Package?	х			
1) FB/EB result non-detect?		х		See table
ICP Interference Check Sample (ICS) included in Lab Package?			х	N/A for Limited Validation
1) Analyzed at beginning of analytical run?			х	
2) %R criteria met? (80-120%)			Х	
3) Spot check accuracy of %Rs			Х	
Matrix Spike/Matrix Spike Duplicate Data Included in Lab Package?	x			JC57819-1A
1) MS/MSD %R (75-125%R) and RPD (20%) criteria met?	х			
2) Was a sample spiked at the frequency of 1/batch or 20 samples?	x			
3) Was the MS performed on a site sample?	х			
4) Was the MS performed on a FB/EB or TB?		х		
Post Digestion Spike			х	
1) %R criteria met? (75-125%R)			х	
2) Was the spike performed on a FB/EB or TB?			х	
3) Was a sample spiked at the frequency of 1/batch or 20 samples?			x	
Laboratory Duplicate Data Included in Lab Package?		х		
Aqueous – Is RPD ≤ 20%?			х	
Soil – Is RPD ≤ 35%?			х	
Was a Laboratory Control Sample (LCS) Included in Lab Package?	x			
1) LCS %R criteria met? (80-120%R).	х			
2) Was an LCS analyzed at the frequency of 1/batch or 20 samples?	х			
Serial Dilution			х	N/A for Limited Validation
1) %D (<10%R) criteria met? -			х	
2) Was the frequency 1/batch or 20 samples?			х	
3) Was a site sample used?	Х			
4) Was a FB/EB or TB used?			Х	

AECOM DATA VALIDATION REPORT FORM – METALS ANALYSIS Page 4 of 5

5) Spot check accuracy of %Ds.		х	
Field Duplicate Data included in Lab Package?	х		JC57819-1A & -2A
Aqueous - Is RPD is <20%?	х		All results met QC limits or were Both ND.
Soil - Is RPD is <35%?		Х	
Percent Solids data included in Lab Package?		Х	Aqueous samples
1) %Solids criteria (Reg 2 criteria) met? (>/=50%)		Х	
Total Cr vs Cr6 Fraction Agreement Reviewed	х		Total Cr greater than total hexavalent chromium.

AECOM Page 5 of 5

Field Blanks

Analyte	Result	3X	10X	Action	Associated Samples
Nickel	4.9	14.7	49	ND, OK	156-MW8B-20171221
					156-MW8A-20171221,
Nickel	4.9	14.7	49	<3X FB, Negate (UJ)	156-MW8A-20171221X

Reporting Limits GW Metals

Sample ID	Lab ID	Analyte	Result	Detect Flag	Units	GWQS Action Level
156-MW8A-20171221	JC57819-1A	THALLIUM	8.2	N	ug/l	2
156-MW8A-20171221X	JC57819-2A	THALLIUM	8.2	N	ug/l	2